

Design of a system to study the adhesive properties of nanoscopic ice layers at cryogenic temperatures

Master thesis by Linnea Widmayer
Date of submission: September 25, 2023

Supervisor: Prof. Thomas Burg
Reviewer: Niko Faul
Darmstadt



TECHNISCHE
UNIVERSITÄT
DARMSTADT

Electrical Engineering and
Information Technology
Department
IMNS

Erklärung zur Abschlussarbeit

gemäß § 22 Abs. 7 und § 23 Abs. 7 APB der TU Darmstadt

Hiermit versichere ich, Linnea Widmayer, die vorliegende Masterarbeit ohne Hilfe Dritter und nur mit den angegebenen Quellen und Hilfsmitteln angefertigt zu haben. Alle Stellen, die Quellen entnommen wurden, sind als solche kenntlich gemacht worden. Diese Arbeit hat in gleicher oder ähnlicher Form noch keiner Prüfungsbehörde vorgelegen.

Mir ist bekannt, dass im Fall eines Plagiats (§ 38 Abs. 2 APB) ein Täuschungsversuch vorliegt, der dazu führt, dass die Arbeit mit 5,0 bewertet und damit ein Prüfungsversuch verbraucht wird. Abschlussarbeiten dürfen nur einmal wiederholt werden.

Bei der abgegebenen Thesis stimmen die schriftliche und die zur Archivierung eingereichte elektronische Fassung gemäß § 23 Abs. 7 APB überein.

Bei einer Thesis des Fachbereichs Architektur entspricht die eingereichte elektronische Fassung dem vorgestellten Modell und den vorgelegten Plänen.

Darmstadt, 25. September 2023

L. Widmayer

Acknowledgment

Ich danke euch

Abstract

Diese Kurzfassung ist kurz.

Contents

1	Introduction	6
1.1	Task and requirements	7
1.2	State-of-the-art	8
2	Experiment preparation and process	9
2.1	Separating the ice layer mechanically	9
2.1.1	Assemblies used at cryogenic temperatures	9
2.1.2	Process	15
2.1.3	Examined factors influencing separation with finger	15
2.1.4	ice structure	17
2.2	PDMS	17
2.2.1	Plasma surface treatment of PDMS	18
2.2.2	Preparation of PDMS samples	19
2.2.3	detaching ice from PDMS	21
2.3	Phospholipids	21
2.3.1	Parylene	22
2.3.2	Preparation of lipid coated slides	22
2.3.3	Solubility lipids	22
3	Experiment execution and analysis	25
3.1	Solubility lipids	25
3.2	Detaching ice mechanically	27
3.2.1	PDMS experiments at room temperature	27
3.2.2	Volume of Hydrofluorether	31
3.2.3	Temperature over applied force	32
3.2.4	Applied Tensile force to HFE	33
3.2.5	Tensile mode vs Shear mode	34
3.2.6	Detaching ice structure	34
3.2.7	observed error sources	40
4	Conclusion and Outlook	41
5	References	43

1 Introduction

Light microscopy is the simplest and most commonly used to magnify small objects and structures. There are many building varieties for different applications with different magnification ranges. In general, two building variation exist: Transmission light microscopy lets light pass through the sample to create an image. Reflective light microscopy uses the reflecting light from the sample to image the sample.

For example, some light microscopes have additional fluorescence filters to detect fluorescence signals. These filters let through only a specific wavelength which is emitted by the fluorescent molecules. As a light source, filtered white light and lasers can be used as only a specific wavelength range excites the fluorescent molecules.

In light microscopy many new advancements are still made. The Abbe criterium is one limit for many light microscopes. However, multiple modified methods exist which are able to resolve beyond this limit [1].

Electron microscopy is another imaging method used. With electrons, resolutions down to atoms are possible. The smallest structures are observable with this method. Even the distribution of certain elements is detectable. Also the creation of 3D images is possible. (MEHR VT LISTEN)

For Electron microscopy, water containing structures must be fixed. One way to achieve fixation of the specimen is to freeze the water to ice at cryogenic temperatures. With rapid cooling, vitrified ice which is amorphous and does not destroy structures can be reached. To hold the ice structure, all setups including microscopes to handle the specimen needs to be cooled below -120°C to prevent the crystallization of the ice.

Conveniently, the process of sample preparation for cryo light- and electron microscopy is very similar. Still, a major difference in the preparation process exists. To reach transparency of the electron beam for cryo scanning electron microscopy (cryo-TEM), the sample is required to be under 100 nm thick and only contains lighter elements. Additionally, a thin film in combination of a mesh is used as a slide. This grid does not deliver a regular background. Additionally, the sample needs low absorption of light in the visible spectrum if a laser is used. With too high contrast, the sample is damaged through the absorbed light and resulting heat.

The sample holder used for light microscopy and electron microscopy are very limited interoperable. This grid does not deliver a regular background for reflective electron microscopy. Also a grid is not completely transparent for light for transmissive light microscopy. Also the grid absorbs visible light, limiting the use of lasers. The slides used in light microscopy are mostly too thick for electron microscopy. Sometimes heavier elements are used too. Additionally some form of metal grid must be added for electron microscopy.

Still, combining Light- and electron microscopy on the same sample would bring big advantages. The larger scale of light microscopy and the usage of multiple wavelengths with fluorescence and high resolution on the same sample will make studying samples easier. In this master thesis, a method for using cryo light microscopy and cryo electron microscopy on the same sample is searched. To make this possible, the sample must be transferred to different sample holders between cryo light microscopy and cryo electron microscopy without destroying the sample. To achieve this, a layer between ice and sample holder is engineered to enable detachment and transfer to another sample holder (fig. 1.1).




Figure 1.1: Depiction of a sample. The specimen is frozen inside the ice layer. The sample holder and ice layer are connected by an additional engineered layer. In this master thesis, the additional layer is varied to allow separation of the ice including the specimen from the sample holder.

1.1 Task and requirements

DAS MUSS ÜBERARBEITET WERDEN

In sample preparation for cryo microscopy, specimens such as cells are frozen inside a thin ice layer. The specimen can be stained with fluorescein before freezing for later observation with cryo light microscopy. Also a sample can be prepared to study with cryo-transmission electron microscopy (cryo-TEM). cryo-TEM allows us to see samples in an hydrated state. This is only possible in cryo-TEM, as liquid water would evaporate in vaccuum [2].

for cryo light microscopy and cryo-TEM, plunge-freezing is used in sample preparation [2] [3]. This can be done either by hand or with a plunge-freezer. Generally the same result can be produced by hand as with a plunge freezer. In practice a plunge-freezer gives more consistent results.

To successfully plunge-freeze a sample, following steps are taken: First, the slide is held by tweezers. Then a 2 mL water drop containing the specimen is pipetted onto the hydrophilic slide. The water droplet is blotted with filter paper, creating a thin film of water which evaporates quickly. The tweezers holding the slide is shot in cold liquid under -140°C , typically liquid ethane. The rapid temperature drop freezes the water into a thin vitrified layer of ice. Vitrified ice has no crystal structure.

A vitrified sample is needed as ice crystals damage the specimen. Vitrified ice is created by cooling water abruptly to temperatures below -120°C [4]. To archive rapid cooling, the slide needs a high thermal conductivity to freeze the water quickly. For example, thin sapphire slides or a metal grid with a film are commonly used. Also the liquid which is used to freeze the sample should not possess the Leidenfrost effect, which prevents instant contact of the sample with the cold liquid. As liquid nitrogen is possessing the Leidenfrost effect, other coolants like liquid ethane are used. Additionall

Currently, no slide exist which is suitable for plunge-freezing, cryo light microscopy and cryo-TEM. in plunge-freezing, a hydrophile surface required to achieve a thin ice layer. Additionally the thermal conductivity of the slide needs to be high for the steep temperature drop needed to create vitificed ice. In cryo-TEM, the sample needs to be extremely thin and small. A metal mesh is used to hold the sample in place. Additionally, only light elements should be used in samples as heavier electrons will disturb the results in cryo-TEM. The mesh used in cryo-TEM is not suitable. The mesh is rough and inhomogeneous with big changes of reflectiveness for a reflected light microscope. for light transmitted through the sample, the light can only penetrate the sample through the holes. This will heavily influence the image quality in cryo light microscopy.

To perform cryo-light microscopy and cryo-TEM in high quality, a sample transfer from one slide to another is proposed. The slide change must be performed at -140°C to maintain the vitrified state of the sample. Additionally as the first slide used for plunge freezing needs to be hydrophilic, which increases ice adhesion in general. First, I investigated lipids for positive characteristics for usage as a sacrificial layer. Second, different layers are tested for low ice adhesion. Also different parameters on the mechanical setup are tested.

1.2 State-of-the-art

Ice removal is needed in multiple commercial fields. But most strategies are applied in temperatures down to -30°C . Since the ice layer in my application needs to stay vitrified at under -140°C or lower, only few anti-frosting methods are feasible. For example the active anti-frosting strategy of heating the ice is not possible as the specimen must stay contained within the ice layer.

There are four passive anti-frosting strategies: Inhibition of ice nucleation is achieved by using surface inherent properties to prevent ice crystals from forming. Retardation of frosting removes water to prevent icing on the surface with water repellent properties such as the lotus effect. Mitigation of frost accumulation prevents already formed ice droplets from further accumulating and forming an ice layer. Last a reduction of ice adhesion on the surface prevents ice droplets to stay on the surface. Other forces like wind or gravity removes ice droplets and keeps the surface ice free [5].

Out of the four passive anti-frosting strategies, only one is applicable at cryogenic temperatures. Inhibition of ice nucleation, retardation of frosting and mitigation of accumulation only inhibit the freezing process itself. Still, the reduction of ice adhesion can be used to decrease the force needed to detach the ice layer mechanically. This can be done by using a hydrophobic surface with general low adhesion. In industrial application, one commonly used coating is Polydimethylsiloxane (PDMS).

PDMS is a polymer which is widely used in different applications like in fabrication of microchannels, chip manufacturing, aerospace industry and medical tools. PDMS properties are for example its hydrophobicity, biocompatibility and electric insulating capabilities. Also PDMS is cost effective and allows rapid prototyping, molding and thin coatings. [6]. Additionally, PDMS is modifiable with additives.

One example in which the low ice adhesion potential of PDMS is illustrated is the passive deicing of Aircrafts in flight. As ice can influence the air flow around the wing an the body, which induces turbulence and reduces lift. Ice protection is therefore critical for a save and stable flying. In [7], PDMS is tuned for optimal characteristics in flight. To test the surfaces, flight conditions of 0.5 bar and -12°C are simulated. Fluorinated PDMS with and without silica nanoparticles are compared to aluminum, showing better resistance against ice growth. The different coatings are also examined regarding contact angle of water and surface roughness. Also the stability of the surface is relevant as ice formation and impacts can also wear down the coating itself.

There are multiple factors which influence the adhesion of Ice on a surface. Hydrophility and the contact angle of water on a surface is for example closely related to ice adhesion. This suggest that plunge freezed samples, which requires a certain hyrdophility, are hard to detach. The adhesion is also dependent on temperature. Experiments performed close to freezing point may not be transferrable to the same setup at -140°C . Also the bulk material and the elasticity of ice itself changes over temperature [8].

2 Experiment preparation and process

To detach the ice layer from the sample holder, two main ideas are proposed. The first method is using mechanical force to separate the ice layer. The second idea is to design a sacrificial layer which can be dissolved at cryogenic temperatures.

2.1 Separating the ice layer mechanically

The goal is to mechanically lift a piece or the whole ice layer from the sample holder it is frozen to. In the whole process, the ice layer must stay in a vitrified state. The sample is frozen with plunge-freezing. For later steps the sample is prepared in a bath. The "finger" assembly is used to attach and pull on the ice layer with Hydrofluorether (HFE). The goal is to vary the engineered layer do reduce the adhesion between ice and sample holder. Also the microscope is used to take pictures of the sample before and after using the finger without heating up the sample.

2.1.1 Assemblies used at cryogenic temperatures

Plunge-freezer, Con

The finger is made of two main parts. The first part is metal rod with a flat tip (Fig. 2.1). The rod is cooled with cold nitrogen gas. Near the tip, the rod is temperature controlled with a Pt1000 used as a temperature sensor and a heater. The second main part is a 3D printed structure, containing the outer layer and routing of the cold gaseous nitrogen. Inside the 3D printed structure, the cold nitrogen gas is directed from the inlet downwards around the metal bar in an inner mantle for cooling. Then, the cold gas is redirected upwards flowing through an outer mantle for additional cooling. Afterwards, the gas exits through the output.

The Cold nitrogen gas is supplied by a liquid nitrogen tank. Strong heaters placed inside the tank evaporate the liquid nitrogen. The volume is rapidly expanding at evaporation, resulting in fast gas flow. Cold nitrogen gas leaving the tank is routed by 6 mm pneumatic tubing (PRÜFEN) to the finger inlet. On the inlet and outlet of the finger, Festo connectors are mounted to allow easy dis- and re-connection of tubes. Cold nitrogen gas which passed through the finger are exhausted by outlet tubing into the atmosphere.

The finger is mounted on three stages. The three stages allow fine adjustment of the finger position in X,Y and Z axis (fig. 2.2). Also, when the finger is attached to a surface, force can be applied by moving the stages in either direction. Additionally the stages are mounted on a track. The track allows fast movement along one axis to make assemblies below the finger reachable. Whenever the finger is used, the assembly is clamped down on the track to prevent additional movement. Whenever the finger is not in use, the assembly can be moved on the track to allow easy access of the area below the finger.




Figure 2.1: A cross-section of the "finger" assembly. The metal rod is cooled with cold nitrogen gas. the gas is routed from the inlet around the metal bar onto an outer layer to the outlet. The metal rod is temperature controlled by a Pt1000 temperature sensor and a heater. With HFE 7200 is applied to the tip, the finger can attach to a surface at cryogenic temperatures and apply force.

On the tip of the finger, Hydrofluorether (here HFE 7200) is applied. HFE is an oil typically used as an cryoimmersion fluid [9]. HFE has temperature dependent characteristics. At freezing point, HFE reaches a viscous state before reaching a firm solid state. This temperature dependency allows to first apply the HFE at higher temperatures with low viscosity and pull on the sample at lower temperatures with high viscosity.

A small bath and later a big bath are placed below the finger. In the baths, samples are prepared and holds a harbor-shuttle system. The sample is fixed in the shuttle for transportation and to hold on to the sample holder when pulling with the finger. A warm nitrogen gas barrier keeps the finger tip ice free by displacing humid air.

In the beginning a smaller bath is used (Fig. 2.3). The small bath contains an elevated floor utilizing as work surface. Embedded in the work surface are indents which fit container holder. These container allow transport and long term storage of samples. Three elevated baths are installed on the elevated floor. They are temperature controlled for containing other Liquids or tools at different temperatures. Also a Haven for a shuttle system is installed. The small baths and the haven are separated from the elevated floor with an insulating layer. The bath is filled with liquid nitrogen covering the work surface. The level is kept on the height of the insulated layer to allow temperature controlling over -195.8°C . The whole bath is insulated by Nitrogen gas flowing inside the 3D-printed shell of the bath. The warm nitrogen gas is expelled from the brim, flowing from the outer edge radially to the middle rotation axis. In that way, the nitrogen gas separates the damp air in the room with the dry nitrogen gas inside the bath.

The usage of the small bath in combination of the finger has limitations: First, the space in the small bath is small. The movement along the track is very limited. The finger can be moved along the track, but the space left still limits work with pincers. Additionally, the smaller baths are not needed when using the finger, therefore taking up much needed space. Also the Shuttle needs to be tilted in a specific angle to dock and undock the shuttle. The work flow also allows only one shuttle at once, limiting throughput. Also liquid nitrogen needs to be refilled often since the bath can only hold a smaller volume.




Figure 2.2: Stages to precisely manipulate the finger tip in X, Y and Z direction. When the finger is not in use, it is moved along the track to allow access to assemblies below.

During this master thesis, a second bigger bath is build (Fig. 2.4). In general, the structure is similar. It also uses an elevated floor as a work surface. The work surface is fabricated out of two plates screwed together and fixed to the brim with 3D-Printed hooks. Indents are formed by holes in the upper plate. No baths are installed, but the space is reserved for later addition. Two harbors are mountable for parallel work on two separate shuttles. The harbors are screwed on an aluminum block with Pt-100 temperature sensors and heaters for temperature regulation. Between the aluminum block and the work surface, a 3D-printed insulating spacer is placed. Also both harbors can be mounted either flat or in an angle, depending of the 3D printed spacer. The bath is insulated with styrofoam and a rim with holes for warm nitrogen gas is placed on top. The holes are places along the inside of the long perimeter.

To take detailed pictures of the sample, an inverted microscope is modified for cryo- microscopy (Fig. 2.5). A box with a metal core is fixed to the bottom of the stage. An outer frame surrounding the box is temperature regulated to prevent buildup of ice. The box is supplied with cold nitrogen gas to cool the metal core. A haven is fixed to the metal core and placed over the objective. The haven is temperature controlled by a ring heater to -140°C . This makes HFE liquid and distinguishable from the frozen sample. to keep the light path between haven and objective clear from ice and fog, a component referred to as "glasses" is used(HIER REFERENZ ZUR ZEICHNUNG VON DER BRILLE?). It contains two thin and parallel glass panes. Warm nitrogen gas is routed between the glass panes and out between the lower glass pane and objective. Additionally, a filter is used for examining fluorescent samples.




Figure 2.3: Small bath used for sample preparation. This bath is combined with the finger. A major drawback of use with the finger is the lack of space inside the bath. To solve this issue, a bigger bath is constructed.

The shuttle allows easy transportation and access to the sample. The sample is clamped down by the "window" brace (fig. 2.6a). The window brace is fixed with two screws to the shuttle, holding the sample in place. A center hole in the window gives access to the top side of the sample for microscopy and finger. In the copper shuttle a thread is placed. With a long threaded metal rod the sample is transported between assemblies.

For long time storage small containers with space for three $\varnothing 5$ mm samples are designed (Fig. 2.6b). These are 3D Printed, modified version of other containers. a cap which is identical with the other versions of container is screwed on top with a special tool. Inside the bath, the container fit into the indent. The cap is screwed next to the container, clamping the container down to the work surface. Also the container are interoperable with most other tools. This allows long term storage of samples with 5 mm diameter.






Figure 2.4: Big bath with finger assembly. Inside the bath samples are prepared. When applying forces with the finger, the sample fixed on a shuttle is held in place by the harbor.



Figure 2.5: Modified microscope for cryo microscopy.



(a) Shuttle for transporting a sample between bath and microscope. The window is clamped down by screws to the copper shuttle, holding the sample in place. The center hole in the window gives access for microscopy and finger.



(b) Small 3D-printed transport box for three samples. A cap is screwed onto the central thread. These boxes fit into storage units. The design is based on a commonly used storing system.

2.1.2 Process

To try out the detachment with the finger in a repeatable manner, a core process is established for orientation. The sample is prepared by plunge-freezing and put in the bath. The copper shuttle part is placed in the harbor. With HFE, the sample is placed in the middle of the shuttle. The HFE helps the sample to stay in place before fixation. The window brace is placed on top and screwed down. The now prepared shuttle is transported quickly to the microscope. When the transfer is not possible within seconds, the shuttle is placed in a portable container with liquid nitrogen. After microscopy, the sample is placed into the bath under the cooled finger.

First, if not already done, cool down the "finger" to -140°C . HFE is applied to the tip. The "finger" is lowered onto the sample while correcting the position with the stages until the HFE contacts and spreads over the sample. The temperature is reduced to -160°C and waited until the sample and finger is cooled down. When the temperature is reached the finger is pulled up by turning the stage until detachment. Then the shuttle is transported to the microscope and the sample is analyzed.

When detachment is successful, the ice piece hanging on the "finger" is placed on another shuttle. This is done by lowering the finger onto the new shuttle and raising the Temperature to -140°C .

To collect first insight, samples with parylene and lipids (as described in Section 2.3) are used first. different variables are determined which could significantly influence successful detaching. Then the different variables are examined with experiments to improve the reliability of the "finger". The different variables found in this thesis are discussed in the following sections.

2.1.3 Examined factors influencing separation with finger

Volume of Hydrofluorether on "finger"

Using the correct amount of HFE is important for higher repeatability. Too little HFE will not bind to the finger and sample. too much HFE results either in a thicker layer, or the HFE flows between "window" and sample. as the cohesion of the HFE is considerably lower than the sample and the finger, a thick layer is prone to breaking before the ice layer. HFE under the "window" clamp will redirect a part of the force, making the sample more stable. This makes detaching also harder.

In the beginning, the HFE was applied with a pincer. The HFE is given in a cold bath at -140°C . HFE gets more viscous at colder temperatures and is evaporating much slower. The thickened HFE is now scooped with pincers onto the tip of the "finger". This method is not used later, as measuring the volume of HFE when applying is not possible.

As an effort to determine the volume of HFE a pipette is used. The HFE is pipetted at room temperature onto the cold finger. When HFE is applied onto the desired surface, around $4\ \mu\text{L}$ has already evaporated. Also, only HFE on the flat surface facing the sample is usable.

Another way to determine the Volume of HFE on the "finger" is by analyzing pictures of the "finger" with HFE. The volume is calculated with the contact angle on the "finger" and the area covered in HFE. With the knowledge if the HFE spreads too much or not being attached properly to the sample, a range can be given where success is more likely. Still, other factors like temperature or the gap between sample and "finger" can influence the result.

Temperature

Ethoxynonafluorobutane, also called HFE 7200, is used throughout all experiments. It has a freezing point of -138°C . Below the freezing point, HFE gets increasingly viscous. at a certain point, the hfe gets brittle and cracks start to form by decreasing temperature. Between the freezing point and the point of cracking, HFE is usable as glue. With a temperature near the freezing point, HFE can spread over the surface. At lower temperatures, HFE gets hard enough to hold the finger to the upper ice layer.

The temperature regulation of the finger has three modes: First the "unglue" mode which regulates the Temperature to -140°C . HFE has a low viscosity, which allows the application of HFE. Also detachment without transferring force is possible. Second, in "glue" mode the shuttle and the finger are cooled to -160°C . HFE hardens and force can be applied to detach the ice layer. Third, "thaw" cleans the finger by heating the tip to 20°C , evaporating everything stuck on the "finger".

As HFE is getting harder with temperature, lower temperatures in the "glue" state could allow higher forces on the ice layer. Still, the point where HFE starts to crack will decrease the tensile strength. To test this hypothesis, HFE is examined at temperatures until -170°C .

Direction of force

The direction of force can increase the likelihood of detachment. Tensile mode and shear mode can be applied by the finger. Tensile mode is the easiest to apply with the finger. Also HFE is able to withstand some tensile forces. But for separating layers, this mode could take more force than applying tensile forces, depending on the bottom layer. Still, HFE stability to shear stress is not known. Additionally, the sample is clamped down to the top. This does not allow the ice surface to slide off without breaking.

In some experiments, the shuttle is tilted around 15 degrees for easier access to the shuttle. The finger is also tilted so the tip surface is parallel to the sample. To apply force, the finger is pulled by stages in either X, Y, or Z direction. For each direction, the stress is split into tensile and shear stress. In Z direction, mostly tensile stress is applied (Fig. 2.7 (a)). in X direction, mostly shear stress is applied (Fig. 2.7 (b)). In Y direction, only shear stress is applied, but will result in shattering of the ice layer, probably taking additional force. For this reason, only X and Z direction are tested.




Figure 2.7: Tensile vs Shear mode

2.1.4 ice structure

DER TEIL BRAUCHT NOCH ARBEIT!!!!!!

To lift off a piece of the ice layer, the ice layer must be broken in some way. thicker ice layers are expected to be harder to break than thinner ones due to the bigger cross section. Also amorphic vitrified ice is expected to be more stable than crystallized ice.

Initially, to save time for experiments, the samples are freezed by Hand in liquid nitrogen, as described before. However, the ice layers are less consistent compared to plunge freezing, resulting in mostly thicker ice layers compared to plunge freezing. also as the sample is frozen in liquid nitrogen, the leidenfrost effect is inhibiting the formation of vitrified ice.

In experiments, the used glass slide and freezing method does not produce vitrified ice. Therefore the influence of vitrification is not observed. The thickness of the ice is also not measured directly.

To compare the influence of hand freezing and plunge freezing, results of lifting off samples frozen with both methods are compared. No other factors are varied in those experiments. In the end, hand freezing and plunge freezing did not make a difference. Therefore, hand freezing was also applied in future experiments, as this effect is determined as neglegtable compared to other factors.

SAMPLE INTEGRITY

positioning

Before attaching the "finger" to the sample, the "finger" is positioned with the stages. At positioning, two different errors can occur: first, the finger is incorrectly positioned over the sample. the HFE is partially on the "window" brace. The second error is a too small or too big gap between "finger" and sample.

An incorrectly positioned finger is expected to be a major error. Depending on the overlap, the force is transferred mostly to the brace than to the sample. This will lead to an unsucessful detachment. With the setup of three stages, the positioning is easily corrected. Alternatively, a "window" brace with a bigger central hole makes positioning easier.

The gap between sample and "finger" is sometimes hard to estimate. The HFE volume can influence the preferred gap size to avoid pressing the HFE between sample and "window". On the other hand, HFE is shrinking between temperatures used at attaching and pulling.

(HIER WEITERSCHREIBEN ODER BEI ERGEBNISSE???)

2.2 PDMS

PDMS is a polymer used in coatings for passive deicing. It is hydrophobic and has a low surface energy. Also it can be coat spinned into a thin layer to form the thin coat. Also it is widely available and tunable. To use PDMS in plunge freezing, the PDMS is plasma activated. The requires a Plasma generator.

To create different PDMS mixtures, Dowsil Sylgard 184 Silicone elastomer is used[10]. The PDMS kit has two components. The base coat component is highly viscous, whereas the curing agent is liquid. The Specified mixture ratio is 10 base coat to 1 curing agent in weight (10:1). In some applications, other mixture ratios are

used and additives are added for tuning PDMS. In my research, I focussed on tuning the mixture ratio of base coat to curing agent.

PDMS properties are also temperature dependent. In [11], multiple characteristics are determined for cryogenic temperatures. The PDMS is prepared with the standard mixture ratio of 10:1. The compressive strength increases with lower temperatures until -123.15°C . At this temperature, the compressive strength reaches a maximum of 224.50 MPa in average. At lower temperatures, the PDMS gets brittle. At -150.15°C PDMS has a compressive strength of 106.99 MPa.

The "finger" requires a temperature of -160°C . Therefore, the temperature drops enough to make PDMS of 10:1 mixture ratio brittle. Generally, a brittle PDMS surface helps to detach the ice layer. Tensile forces loosens the PDMS under the Ice. The PDMS layer breaks within itself, loosening the ice layer from the rest of the sample. Still, an extreme brittleness is needed, which may not be reached. With PDMS of other mixture ratios, the brittleness could change, including the temperature dependency.

2.2.1 Plasma surface treatment of PDMS

Plasma curing is commonly used as PDMS treatment. For example, it is used to bond two PDMS surfaces together [12]. It is also used for increasing wettability and adhesion. Plasma treatment is changing the chemistry of the polymer chains on the surface. Charged Oxygen Ions are deposited on the surface. these Ions make the surface temporarily hydrophile and increasing water adhesion. The Ions change the structure of the PDMS is permanently by oxidation. In some cases, cracks form as the surface oxidizes to a silica like form.

In [13], the influence of different gasses on PDMS plasma treatment is examined. Oxygen, Nitrogen, Argon and Helium are compared on the effect on wettability, adhesion and cracking. Thin PDMS sheets are used with unknown composition. They found similar results between gasses. All gasses produced a thin and brittle surface with cracks and high wettability. Based on these results, used gas is not a significant factor for plasma activation. Therefore using only air (mainly a mixture of nitrogen and oxygen) is sufficient to determine the effect of plasma treatment.

In [14], the influence of plasma treatment on Mixture ratios of 50:1 to 100:1 is described. A thin layer of those PDMS mixtures is put onto a preformed PDMS piece with lower mixture ratio. The preformed piece is used to apply shear stress to the surface by stretching the lower PDMS form piece. Therefore only tensile force are examined. It shows no significant difference between described mixture ratios. It also shows that higher plasma treatment leads to more brittle surfaces, reducing the force needed to break the PDMS Layer by 90%.

Also known are the adhesion forces on PDMS from 1:3 to 50:1 mixture ratio. In [15], PDMS is mixed in different weight ratios. The mixture is given into a mold and afterwards fully cured. An ice block is frozen on top of the PDMS. Using a pulling machine the maximum tensile and shear forces for detaching ice from PDMS are determined. Results show that mixture ratios 10:1 to 1:3 have significant lower adhesion forces on ice. At for example at 2:1, the shear mode is just under 20 kPa and the tensile mode around 30 kPa.

Still, the effect of plasma activation under 50:1 is unknown. The plasma activation has two effects: first, it can increase the adhesion on ice to PDMS. second, the PDMS changes its structure, which could increase as well as decrease the strength of the PDMS layer. In the following, the effect of plasma activation on PDMS are examined. This is done at room temperature to speed up the process.

2.2.2 Preparation of PDMS samples

All PDMS samples are prepared in a similar way. The preparation starts with weighting out the desired amount of base coat and curing agent. The mixture is stirred intensively. The mixture is placed under a vacuum bell to gas out all air bubbles from stirring. Meanwhile the glass covers are cleaned with ethanol or isopropanol and dried. Afterwards, the PDMS mixture is coat-spun onto the glass covers. Then the coated cover glasses are baked in the oven.

For 1:2 base coat to curing agent weight ratio, the PDMS mixture is lightly viscous. A vacuum of 30 min is drawn for degassing. A coat spinning time of 120 min at 3000 RPM results in a smooth surface on all used slides. The baking time of at least 24 h by 80 °C is needed. For shorter baking times, plasma treatment has a slightly different effect. Normally, touching a treated area will neutralize the effect of plasma treatment like hydrophilicity only locally on the touched surface. But here, touching the surface leads to the complete neutralization of plasma treatment when touching. In this work, the effect is undesired.

For 4:1 base coat to curing agent weight ratio, the PDMS mixture is more viscous. The Vacuum and coat spinning are the same for 30 min vacuum and 3000 RPM for 120 min. But a baking time of 30 min at 80 °C is already sufficient to harden the PDMS.

For 50:1 base coat to curing agent weight ratio, the PDMS mixture is as viscous as the base coat. A longer vacuum of 1 h is needed to air out all bubbles. The coat spinning speed and time is increased to 3500 RPM for 180 min (NACHSCHAUEN). Also a longer baking time of 20 h at 80 °C is needed to harden the PDMS.

In the coat spinning setup, each slide is coat spun in succession of each other. Each coat spinning process needs 2 to 3 minutes. This results in a time consuming process. To speed up the process, rectangular cover glass with 20x20 and 24x40 are coat spun. afterwards the cover glass is split in multiple smaller pieces.

The process is the same described as before for each PDMS mixture ratio, but with some adjustments. Before coat spinning, the glass is scratched with a diamond pencil. The scratches in the glass are weak points for breaking out smaller parts. Tape is put over the scratched glass. The glass is put into the coat spinner with the taped side facing down. The PDMS is coat spun onto the exposed glass side. The baking process is as previously described. After baking, the glass is broken into smaller pieces. This is done by dragging the attached tape over a table edge. Then the glass is fixed to the table with PDMS facing up. The PDMS is covered with (????) to keep the PDMS clean. As the sticky side of the tape is not facing the table, the tape ends are additionally taped to the Table. Then smaller sample pieces can be broken off. The pieces are loosened from the tape by forcing a flat pincer between glass and the tape it is attached to.

This method has advantages as well as drawbacks to coat spinning small glass pieces separately. First, time is won by coat spinning, as one big glass can be split into several smaller ones. The Sample can stay fixed at the table without risking damaged. The samples can be broken off shortly before the experiment. But with hand scratching, only irregular and rectangular shapes can be won out of the bigger glass piece. also there is a significant loss of samples. This is by one part by breaking the sample. Some cracks do not follow the scratch, which leads to too small samples. Also in the process of loosening a piece cracks can form, making the piece too small for use.

In the end, these two different fabrication methods are used for different experiments. The small round cover glasses are used in combination with the finger. The regular pieces are easier set up with the shuttle. The glass pieces broken out of a big cover glass are used under the pulling machine at room temperature. In this setup, fluctuations in size of the glass pieces are easier to handle and the clamp has more flexibility.

Setup of the pulling machine

To test the tensile strength of different PDMS mixtures and the effect of plasmacuring, a Pulling machine inspekt table 5 kN manufactured by Hegewald & Peschke is used. On the top part, two load cells are installed (Fig. 2.8a) The upper load cell is rated for 5 kN. the lower load cell is rated for up to 100 N. As the upper one is extremely stiff and the forces are $\ll 100$ N, the upper sensor is assumed as inflexible. On both the upper and lower part, two clamps are fixed onto the machine. On the bottom clamp a 3D-Printed stage is used for fixing on the sample.




Figure 2.8: Setup for tensile tests of a thin PDMS layer. The Sample is placed on the stage. A brace is screwed over the sample to hold it in place. A stamp is fixed to the top clamp. The UV Glue is applied between stamp and sample and cured with 3 minutes UV Exposure. With the 100 N load cell, the force is measured over distance pulled. The load on the 2 kN load cell is neglectable as the forces measured are $\ll 100$ N.

On the bottom, a 3D-Printed stage is clamped (Fig. 2.8b). The stage has four holes with threads for screws . Between the threads, the sample is placed. A brace similar to the "window" is fixed with screws onto the sample. A rectangle shape hole gives access for the UV-glue and the stamp.

The Process of a pull test starts with clamping a new stamp on the top. Then the sample is plasma treated. After treatment, the sample is quickly transported to the pulling machine and fixed to the stage with screws. The stamp is aligned to the sample. When alignment is finished, the UV glue is given onto the stamp. the top is lowered on the sample. The UV glue is cured with UV radiation for 3 min. After gluing, 3 min are waited until forces settle resulting from curing and heat. Then, the pulling machine pulls with constant speed and measures the stress strain curves. The process is ended manually. The now separated stamp and sample are stored for analysis under a microscope. To calculate the maximum tensile stress, the area of the glue on the sample which is still attached to the stamp is measured.

2.2.3 detaching ice from PDMS

Now, different PDMS mixture ratio with different plasma activations are tested at cryogenic temperatures. With previous setup, plasma activation of 1:2 mixture ratio PDMS resulted into a more stable layer with more adhesion. Also 50:1 is tested with high plasma activation. The tensile stress of 50:1 mixture ratio is 60 kPa [15]. [14] suggests that strong plasma activation over 3 min results of a decrease of adhesion strength of around 90 %. Additionally, 4:1 mixture ratio with high plasma activation is tested.

Two different setups are used. First, the setup with "finger" and "bath" described in section 2.1.2 is used. Second, for additional data, a modified version of the setup with the pulling machine is used. This allows the measurement of the actual tensile strength on the sample.

Setup pulling machine for cryogenic tests

To allow tests at cryogenic temperatures with the pulling machine, the setup is modified to allow fitting a bath and the "finger" to the pulling machine. The big bath is used with a flat harbor. An outer frame is used and modified to hold the bath. The bottom clamp is removed and the Bath is fixed over the bottom attachment. The top part is used with the same load cells, but with a bigger clamp. The "finger" 3D printed shell modified with a flat outer profile for clamping and clamped onto the top. 90° festo tubing is used to connect the cold nitrogen gas supply. A power source and a temperature regulator is again used.

BILD

Compared to the "finger" setup with stages, the finger only moves up and down. The Alignment is done by moving the bath by loosening the screws on the frame. Also new "window" parts are waterjet cut. The new "window" has a bigger central hole for easier alignment.

As the pulling machine is set up in a separate room, the samples are prepared in the laboratory next to the microscope. The small bath is used for sample preparation. The process is the same as all detachment trials with the finger.

2.3 Phospholipids

Phospholipids are the building block of membranes in nature. They are made of two long nonpolar carbon chains and a polar head. A membrane is a bilayer of Phospholipids with the hydrophobic carbon chains oriented inwards and the hydrophilic head pointing outwards. They are also natural detergents, as they can bind to hydrophobic waste forming an emulsion, making them removable with polar liquids like water [16].

Phospholipids are dissolved in some common solvents. To apply Phospholipids, a slide is dipped into the solution containing the phospholipid. When the solvent evaporates, the lipids are binding to the surface of the slide, forming a layer with varying homogeneity. this layer is removable with the same solvents. If the ice layer is frozen onto the lipid layer, the lipids can be dissolved at cryogenic temperatures as a sacrificial layer to detach the ice. But to dissolve this layer, a high solubility at cryogenic temperatures is required. The ice layer on top is a barrier between solvent and solute, reducing the area where solvent and lipids are directly in contact.

2.3.1 Parylene

Parylene is a hydrophobic polymer used as a coating to repel particles, including water and ice. Parylene is also biocompatible and used in medicine and biology [??](#). Parylene is not usable without a second layer on top. Parylene hydrophobicity does not allow water to spread during plunge freezing. with plasma activation, the surface is now hyrdophilic, but ice adheres to the parylene too strong to mechanically detach. also parylene cannot be dissolved with a solvent as a sacrificial layer.

For this reason, lipids are used in combination with parylene (Fig. 2.9). The hydrophobic chains of the lipids adhere to the parylene. The polar head allow water to spread evenly over the surface. Solving the lipids with a solvent will detach the ice layer from the slide. Parylene additionally prevents (re-) attachment through holes in the lipid layer.




Figure 2.9: Layers of a Sample. The Lipid layer is used as a sacrificial layer. To reach the layer with a solvent, the only contact surface is to the edge. To get a fast and reliable process, a solvent with high solubility is needed.

2.3.2 Preparation of lipid coated slides

To create the slides with parylene and lipids, a cover glass (5 mm diameter) is used as base. the cover glass is coated with a thin layer of parylene. The coated cover glass is dipped into lipid solution. The cover glass is dried, leaving behind a lipid layer. The prepared slides are then used in plunge freezing.

Two different kind of lipids are used: DOPC and EGG-PC. DOPC is storaged as a powder. The DOPC powder is solved in Ethanol (25 mg/1 mL lipid to solvent) for application. EGG-PC is shipped solved in chloroform in two different ratios: 25 mg/1 mL and 10 mg/1 mL. The solution is shipped in vials.

The lipid solution is transferred into several small bottles. small bottles are chosen because solution forms a lubrication film on the thread of the lid. this prevents the lid from closing airtight. This leads to evaporation of the solvent over time, making the bottle unusable. In the coating process, solution often drops onto the threads, making a bottle only usable in one coating session. By splitting the solution into multiple flask, more slides can be covered from one batch of solution.

2.3.3 Solubility lipids

Tests are conducted to find a solvent to dissolve a sacrificial layer out of lipids. Two consecutive solubility experiments are proposed. The first experiment is conducted at room temperature. the aim is to find solvents with high solubility at room temperature. the candidates with high solubility are then tested in the next experiment at cryogenic temperature. the aim is now to find solvents with also high solubility at cryogenic temperatures. The first experiment is conducted as there are only three baths available at cryogenic temperature. therefore the throughput for experiments is limited.

at room temperature

The Solvent are chosen based on availability, freezing point and safety. The solvents are all readily available in the laboratory. Some were ordered before the test. Also all chosen solvent are save to use in a well ventilated room. The followup experiment cannot be conducted under a extractor hood as too much space is taken up with the experiment. Also the solvent or solvent mixture needs to stay liquid at around -140°C to assure that the ice layer on top stays vitrified. The tested substances are 4-Methyl Pentene, 3-Methyl Pentene, 1-Pentene, Isopentane, 1-Propanol, Pentane and Ethanol.

Each solvent is put in a separate bottle. The lipid coated slides are prepared as previously described. For each solvent, a slide is put in the corresponding bottle. After 15 min, the slides are removed and examined. The results are documented in a list. When all streaks caused by the lipid layer disappeared, the solvent is tested in the next experiment.

at cryogenic temperature

Solubility is temperature dependent. The process of dissolving a solute in a solvent can be broken up in three parts: First, the structure of the solute must be broken up with energy. Second the structure of the solvent is broken up also taking energy. Last, the solute and solvent are rearranged into a new structure which releases energy. Depending on the ratio of energy taking up by breaking the structure of solute and solvent and energy released at rearranging, the process is endo- or exothermic [17]. Exothermic processes have higher solubility at low temperatures, while Endothermic processes have higher solubility at high temperatures [18]. Therefore, solubility needs to be tested at the same temperature as in application.

A quick look at the chemical structure of lipids reveals that solving lipids is most likely endothermic. Lipids ordered in layers are held together with van-der-vaals forces, hydrogen bondings and electrostatic bonds [19]. This results in a very strong bond between lipids. Therefore, a high amount of energy is needed to break up the lipid layer structure. This makes a resulting endothermic dissolving process likely.

The experiment is conducted at -140°C . The solvents are given in liquid nitrogen cooled baths, which are regulated to the desired temperature. A slide is given into the cold solvent for 15 min. Then the slide is examined for leftover streaks as before.

The freezing point of tested solvents are not all below -140°C (Table 2.1). Still, solvents with a high freezing point can be mixed with other solvents with lower freezing point to lower the freezing point of the mixture. Alternatively, the temperature can be raised over the freezing point, but this could risk the ice to loose the vitrified state.

solvent	melting point in $^{\circ}\text{C}$
4-Methyl Pentene	-154
3-Methyl Pentene	-154
1-Pentene	-165
Isopentane	-160
1-Propanol	-126
Pentane	-129
Ethanol	-114

Table 2.1: Melting Point in $^{\circ}\text{C}$ for tested solvents.

Additionally, some solvents tested are soluble in water. It is unknown whether the solvents could be solved or diffuse inside the ice layer at -140°C . Therefore the ice layer could be changed in some undesired manner. if a sufficient solvent is found and the solvent is soluble in water, a potential change of the vitrified ice needs to be addressed.

3 Experiment execution and analysis

3.1 Solubility lipids

In the previous chapter, the method of using a sacrificial layer to detach ice is discussed. For this, lipids need to be solved at cryogenic temperatures. As not every lipid is solvable in all solvents, an experiment is conducted to obtain solvents at room temperature. Then the best solvents are tested at cryogenic temperatures.

The solubility of lipids at room temperature in different solvents are determined. For this experiment the cover glasses are coated with lipids. Then a first reference image was taken. Then the cover glass is given into a small container with the potential solvent. After 15 minutes, the cover glass is removed and compared under the microscope with the reference picture. If streaks created from lipids are still as visible as before, the lipids are categorized as insoluble in this solvent. If the streaks partially disappeared and/or are less visible, the lipids are categorized as partially soluble in this solvent. Last if the streaks completely disappear, the lipids are assigned as soluble in the solvent (Table 3.1).



Figure 3.1: Example of a Ø5 mm cover glass with lipid residuals on the surface.

This experiment shows that each three different solvent exist for EGG-PC and DOPC with high solubility (Table 3.1). Following these results, solvents categorized with "soluble" are tested regarding solubility at temperatures of -140°C . As not all solutions are liquid at -140°C (Table 2.1), they are tested at higher temperatures above their melting point, as mentioned in chapter 2.3.3. In addition they are tested as mixtures with other solvents with a lower melting point. Additionally liquid ethane is tested as solvent. Ethane was not tested at room temperature, as the boiling point is at -88.6°C [20].

potential solvent	solubility EGG-PC	solubility DOPC
4-Methyl Pentene	soluble	N/A
3-Methyl Pentene	slightly soluble	insoluble
1-Pentene	insoluble	insoluble
Isopentane	soluble	slightly soluble
1-Propanol	soluble	soluble
Pentane	soluble	insoluble
Ethanol	N/A	soluble

Table 3.1: Result of solubility tests at room temperature. Soluble indicates solvents which are able to visibly solve all lipids off a cover glass. slightly soluble indicates solutions which are able to solve lipids with residuals. Insoluble indicates no visible removal of tested lipid.

In the experiment, no tested solvent was able to completely solve lipids at -140°C and within 15 min (Table 3.2). Also streaks of applied lipids did not only stay partially behind, but also new streaks appear on the glass slides. This means that some lipids redistributed on the same glass slide.

Using solvents to remove a sacrificial layer, a high solubility is required. In practice, the sacrificial layer is completely covered by the ice layer except the edges. Therefore area of contact with the solvent is small, slowing the process considerably. Additionally, as the ice layer needs to stay vitrified. The temperature cannot be raised over -140°C to speed up the process.

The solving process of lipids proves to be endothermic. This means that heat is needed to solve lipids, so cold temperature heavily decrease solubility. This effect was observed over the last experiments by all solvents to varying degree. It can be assumed that the majority of solvent-lipid mixtures are endothermic which is very disadvantageous for finding a potential solvent-lipid candidate. Strongly exothermic solvents could heat up the ice enough to crystallize the ice. So weakly exothermic solvents would be optimal for this task.

Solvent	Result
Pentane	soluble at -125°C
4-methyl pentene	insoluble
1:1 volume ratio HFE to 1-Propanol	did not mix, slightly soluble
Liquid ethane	insoluble

(a) EGG-PC

Solvent	Result
1:4 volume ratio 1:2 molar ratio Ethanol to Isopentane	slightly soluble
1:2 volume ratio 1:1 molar ratio 1-Propanol to Isopentane	insoluble
Isopentane	slightly soluble
1-Propanol	at -130°C slightly soluble
Liquid ethane	insoluble

(b) DOPC

Table 3.2: in 3.2a for EGG-PC, no sufficient solubility at -140°C was found. In 3.2a, DOPC was tested but also no proper solution was found.

3.2 Detaching ice mechanically

3.2.1 PDMS experiments at room temperature

To speed up the process of finding the right balance of PDMS Mixture ratio and plasma curing, experiments at room temperature done. For this, a pulling machine is used. First the sample is prepared. Then, the sample is clamped to the pulling machine. Then a plexiglass stamp is aligned on top of the sample. With UV glue, the stamp is glued to the PDMS layer on the sample. Before gluing, The relative force and distance is set to zero on the pulling machine. After gluing, a couple minutes are waited so no further stress changes are ongoing from the gluing process. After setting the force again to zero, the Machine is pulling on the sample with constant velocity. After detachment, the measurement is stopped. Afterwards, the stamp and the layer are analyzed under a microscope. The area of the attached UV glue to the sample is determined. With the maximum force and area, the maximum stress is calculated. Each experiment is repeated multiple times.

In between experiments, small variations and improvements are done: two different stamps are used, one has an area of 2 mm x 3 mm and another stamp is 3 mm x 3 mm. Since the area is measured by the attached surface to the sample, the effect of a bigger stamp is taken into account. Also, in first experiments, while waiting of the stress changes to subside, the pulling machine is holding the position. As the UV glue cools and shrinks while being polymerized, the UV glue pulls the stamp and the surface more and more together. Before pulling, the machine is set back to zero. After pulling, an offset exists between before pulling which is set to zero and after pulling when both ends are disconnected. The offset is then corrected in evaluation. To

avoid calculating an offset, the pulling machine sets the distance corresponding to the forces being constant to zero. The offset correction and new method increased the accuracy between pull tests.

To verify the setup, cover glass coated with 4:1 and 1:2 curing agent to base coat weight ratio and uncoated cover glass as control surface are compared. The results show 2:1 mixture ratio with 87.3 ± 19.9 kPa is easier to detach than 1:4 mixture ratio with 429.1 ± 5.1 kPa (Fig. 3.2). Also the cover glass without PDMS takes up a lot more tensile stress up to 1161.5 ± 111.5 kPa. sometimes the machine is able to break the control cover glass.

In literature, the ice adhesion on PDMS without plasma treatment is 35 kPa. for 2:1 and 5:1 the stress is between 60 to 80 kPa [15]. This is considerably lower than the experiment before. Therefore, one limitation is that the actual adhesion between ice and PDMS cannot be simulated by this experiment. Still, there is a correlation between values measured and in literature. Additionally, A detachment between other layers than the transition of ice and PDMS counts as successful detachment.




Figure 3.2: Comparison of PDMS with 4:1, 1:2 Base coat to curing Agent mixture ratio by weight coated on a cover glass and cover glass without PDMS as reference. the cover glass without pdms is clearly distinguishable by higher tensile stress. A mixture ratio of 1:2 shows the lowest tensile strength.

In the next experiment, the effect of plasma curing is investigated. A mixture ratio of 1:2 is examined as this mixture ratio leads to the lowest adhesion to ice and other surfaces. The same setup on the pulling machine is used. Samples with a 2:1 weight ratio PDMS are additionally plasma treated before quickly clamping on

the pulling machine. The minimum plasma activation possible with used plasma machine is 25% at 0.1 min. Below 25% plasma does not reliably form in the chamber. the maximum tested plasma activation of 2 min by 100% is additionally done with a bigger plasma machine.

After pulling, The UV glue is almost always firmly connected to the Stamp. In some cases, the UV glue is not connected to the Stamp. This suggest that the glue did not adhere well enough to the stamp. These attempts are sorted out. For the same reason, attempts where the cover glass cracked are sorted out as well. Additionally, cases in which the Tensile stress is near or within the uncertainty of the control are sorted out as well. This can occur when the PDMS slide flips between loosening and fastening.

Even with low repetition rates of $n = 1$ to $n = 3$, a clear tendency is visible. With lower and stronger plasma treatment, the tensile stress needed to detach from the PDMS Layer is higher (Fig. 3.3). Over the whole range, The tensile strength of PDMS becomes up to six times higher. This is not expected, as plasma activation on mixture ratios of 50:1 has the opposite effect [14]. This means that this the results are not applicable to other mixture ratios. also no visible changes to a glass-like state was observed in 2:1 weight ratio mixture. Because the repetition rate is low, the exact values should be treated with caution.




Figure 3.3: Tensile strength of PDMS 2:1 mixture ratio coated cover glass with different plasma activation. For each data point, the repetition rate is $n = 1$ to $n = 3$. A clear tendency of an increase of tensile Stress over intensity of plasma activation is visible. TODO PLOT ANPASSEN, VIELLEICHT DEN PUNKT BEI 2 min raushauen

As PDMS is hydrophobic, plasma activation is needed to freeze a thin layer of ice onto the coated slide. However, super low plasma activation could not decrease the contact angle of water enough. To test the contact angle, PDMS coated slides with mixture ratio 1:2 is plasma activated with 25% for 0.1 min, 30% for 0.2 min and 35% or 0.3 min. 25% for 0.1 min results in a contact angle around 45°. 25% for 0.2 min contact angle is around 15°. 35% with 0.3 min is below 10°. As a control, a PDMS coated cover glass with no treatment and a PDMS coated cover glass with UV treatment of 3 min are compared. The untreated and UV treated PDMS coated glass have both a contact angle of over 90°. Both 25% for 0.2 min and 35% with 0.3 min are used in experiments.




Figure 3.4: force over Time

Additionally, two other PDMS mixtures are used. 4:1 with 10 min and 100 % plasma activation is used as reference. 50:1 mixture ratio is additionally used with different plasma activation duration. To verify the results of [14], PDMS with 50:1 mixture treated with 3 min at each 25 % and 100 % and 10 min at 100 %.




Figure 3.5: Comparison of different Plasma activation on PDMS with 50:1 mixture ratio. In (a), 25 % Power at 3 min leads to no crack formation. in (b), 100 % Power at 3 min leads to cracks. Increasing the duration to 10 min at 100 % power increases the cracks. The cracks indicate increased brittleness of the surface.

3.2.2 Volume of Hydrofluorether

First the volume of HFE is evaluated. High dosages of liquid HFE can spread underneath the frame holding the sample, leading to an inefficient force distribution. Also a thick glue layer take less tensile force, leading to a reduction of maximum force on the sample. Too little HFE will not connect the finger to the sample. Additionally, the dosaging of glue ia found to be a challenge.

The HFE is dosaged with a pipette onto the tip of the "finger". While pipetting, around $4 \mu\text{L}$ of HFE is evaporating. Based on this knowledge, dosaging $4.10 \mu\text{l}$, $4.30 \mu\text{l}$ and $4.50 \mu\text{l}$ is compared and a video is taken for later comparison.

The videos show that pipetting HFE is not reliable. The visible HFE volume on the tip does not correlate with the dosaged volume. On reason is a difference in the volume of HFE evaporating while applying. Also some HFE may be placed on the side of the tip. When using the finger, only the HFE on the flat tip is effective.

Still, a visual estimate for the correct glue dosage is made by calculating the drop volume out of camera



(a) lower limit HFE volume



(b) upper limit HFE volume

Figure 3.6: Visual representation of the range of dosages which are the most reliable. less HFE than the lower limit is more likely to not attach to the sample properly. more HFE than the upper limit will spread under the "window" brace or results in a thick HFE layer with less tensile strength.

images. Two exemplary Pictures of an upper and lower limit of glue dosages is picked (Fig. 3.6). Then the volume is calculated with a formula for the volume of a spherical section. All needed components are calculated out of the estimated contact angle of the glue $\alpha \approx 45$ and the tip diameter of $d = 1.68\text{ mm}$. For the lower range a reduction of d by a factor of $\frac{2}{3}$ is assumed as the drop is not covering the whole tip. The resulting volume range of the HFE is $0.11\text{ }\mu\text{l} \gtrless V \gtrless 0.38\text{ }\mu\text{l}$.

3.2.3 Temperature over applied force

The temperature dependent viscosity of HFE is used to attach and detach from the surface of the sample. With decreasing temperature, the viscosity of HFE increases. At some point, HFE forms cracks and gets brittle. To optimize the applied force at detaching, a temperature is needed which gives HFE high viscosity but does not form cracks at the same time. The determined temperature is then set for the "glue" state described before.

The viscosity and crack formation of HFE are tested. for this, HFE is given into a small temperature controlled bath. the temperatures -150°C , -155°C , -160°C , -165°C and -170°C are compared. Also changes in HFE when cooling and heating up are observed. A needle is put in the HFE to induce forces and subjectively test the viscosity of HFE at all temperatures.

At -150°C to -155°C , the HFE is still only lightly viscous . the needle is not held up by the HFEs viscosity. At -160°C to -165°C the HFE is viscous enough so that the Needle is hold up by the HFE. The Needle can be pulled out and the HFE is closing the gap. Also with enough force, the Needle can penetrate the HFE. also no cracks formed so far. At -170°C , HFE hardens further. The HFE is still viscous. Wiggling the Needle in the HFE can result into cracks. Then the Needle can be easily pulled out. Under -170°C , Cracks form without inducing forces in the HFE.

Heating the cracked HFE up leads to the cracks eventually disappearing. at -165°C , first cracks disappear, but a many remain. The cracks left are still lowering the mechanical stability of the HFE. Heating up to -160°C results in less cracks, but some still remain. A temperature of -150°C will result in cracks completely disappearing.

At -165°C , maximum load can be applied. Higher temperatures result in lower viscosity, which lowers the maximum stress before HFE breaks. At -170°C , cracks will form, lowering the maximum stress.

To test how applicable the temperatures are with the "finger" setup, pulling tests are done as described in section 3.2 except the temperatures of the finger is lowered to -165°C and -170°C in the "glue" state. With attention to proper insulation and no leakages of cold nitrogen gas, -165°C is quickly reached and is held stable by temperature regulation over a time span of minutes. Still, the setup is not very reliable as new leaks can form and changing of tanks/tubings can lead to new leaks. These leaks are spotted only when the finger is already cooled. To fix leaks, the setup needs to warm up. -170°C can sometimes be reached with the finger, but holding the temperature stable is not possible.

When cooling down to the desired temperature with the "glue" state, liquid nitrogen is refilled to increase cooling. When refilling and cooling at the same time, rapid cooling happens when liquid nitrogen touches the shuttle directly at refilling. With this, the Temperature can shortly drop below -170°C . This induces cracks in HFE, lowering the tensile strength. If this happens, the Sample and finger can be heaten up to the "unglue" state at -140°C and the cracks disappear. Then cooling can start again.

In conclusion, the finger "glue" temperature is most effective at -165°C . Still, the reliability suffers trough leaks, longer tubing and improper insulation. For better repetition, a temperature of -160°C is also used in experiments.

3.2.4 Applied Tensile force to HFE

To additionally measure the force applied of the finger, the pulling machine is used in combination with the finger. In the lab, the small bath is used for sample preparation. after microscopy, the sample is transported to the pulling machine. the big bath is fixed on the bottom of the pulling machine. on top, the finger is clamped into the shackle and aligned to the shuttle. The process is still as previously stated.

This setup has a lot of limitation. Aligning the finger is very difficult, as the screws are hard to reach in cold temperature. When cooling, the previous alignment ist lost as the coldness distort the setup. Also little movement at the finger has maximum effect on the sample and alignment. touching the tubes can loosen the sample and/or the setup needs to be new aligned. Even a bigger window is not enough. Errors in alignment and therefore gluing to the border of the window and loosening through movement are the two biggest error sources in this setup. So the old setup is used in further experiments to eliminate big error sources.

Over all pulls, the HFE can transfer a maximum force of $1.30 \pm 0.49 \text{ N}$. The area which experienced the force is unknown. Additionally, the force distribution is uneven (e.g. like Fig. 3.7c). Still, a worst case and a best case which are easily calculated can be determined. as best case, all force is evenly distributed under the area of the finger (Fig. 3.7a). With this assumption, the Area is equal to the finger tip area. The worst case is an even distribution over the whole surface (Fig. 3.7b). The surface is clamped by the window. So the inner diameter of the Window is the worst case area. In reality, the Force is somewhere in between (Table 3.3).




Figure 3.7: Different worst- and best case examples of force distribution on sample. In best case, the force is concentrated on the small area around the finger tip (a). In worst case, the force is evenly spread over the whole sample (b). In reality, the force distribution is between shown extremes. One example of a realistic force distribution is shown in (c).

Area used	Area Size	Tensile Stress
Finger	2.217 mm ²	585.4 ± 219.4 kPa
Small Window	4.91 mm ²	264.3 ± 99.1 kPa
Big Window	12.56 mm ²	103.3 ± 38.7 kPa

Table 3.3: Tensile stress and size of area used to calculate the tensile strength of the finger setup. In worst case, the force is distributed over the whole "window" size(Fig. 3.7b). In best case, the force is only distributed on the "finger" tip (Fig. 3.7a). The real value is between those extremes.

3.2.5 Tensile mode vs Shear mode

THIS NEEDS MORE LOVE

As mentioned before, Force can be applied by moving the stage in either X, Y or Z axis. In the tilted position of the harbor, moving in Z axis splits in mostly tensile mode but also shear mode.


Pulling tests are compared along the X and the Z axis. Generally, way less force is transferred when moving along the X axis. this is feelable with the resistance of the stage and hearable when the HFE detaches.

3.2.6 Detaching ice structure


Different ice structures and thickness results in different stability of the ice layer. The freezing process has an influence on the formation of the ice structure. To compare plunge freezing by a plunge-freezer to the process by hand, samples with lipid coated slides which are prepared with a plunge freezer are compared to samples frozen by hand. The water applied to form the ice layer is mixed with fluoresceine. The fluoresceine to water ratio is 50 mL/10 mg and 5 L/10mg, depending on the microscope setup and objectives.

Freezing by hand results in less predictable shaped ice layer. Sometimes, a non continuous layer forms (Fig. 3.8 (a) and (b)). In other iteration, a continuous layer is formed. (Fig. 3.8 (c) and (d)). A plunge freezer reliably produces continuous layer (Fig. 3.9). The thickness between samples varies. Some ice layers are thick enough to include air bubbles. Also, both hand-freezed and plunge-freezed samples ice layer have sometimes visible cracks.


To further compare both methods, 4 of each hand freezed and plunge freezed samples are compared regarding detachability. The "finger" is attached with HFE onto the sample. With movement up on the Z axis, the "finger"




(a) Sample with a broken ice layer in true light



(b) Sample with a broken ice layer with fluorescence filter



(c) Sample with continuous ice layer in true light




(d) Sample with continuous ice layer with fluorescence filter

Figure 3.8: Two different examples for hand freezed ice layers on samples.


is inducing tensile force to the ice layer as an attempt to detach an ice piece. The results are categorized in 4 categories: Not successful pulls don't have visible changes of the florescent ice layer, Partially successes are visible breaks or clear movement of ice parts on the ice layer, Successful liftoff is a missing piece and a visible piece on the finger, which could be used for future steps.

In the results, there is no difference between hand-freezed and plunge-freezed samples regarding detachability (Fig. 3.4). Therefore, using a plunge-freezer does not result in samples with favorable properties.


To compare PDMS coated slides to lipid coated slides, the same experiment is repeated for three different mixture ratios. Under the microscope, a continuous layer of ice is visible. 3.10). The Ice layer has the same structure for each mixture ratio. Still, the ice layer has a different structure to ice frozen on lipid slides (fig




(a) Sample with thin continuous ice layer in true light



(b) Sample with thin continuous ice layer with fluorescence filter



(c) Sample with thick continuous ice layer in true light with air bubbles



(d) Sample with thick continuous ice layer with fluorescence filter with air bubbles

Figure 3.9: Two different examples for hand freezed ice layers on samples.

3.8, 3.9). The layer is only interrupted by the imprint of pincers used for plunge freezing, visible in fig. 3.10f. Also almost no cracks are visible in the ice layer. This indicates a strong continuous ice layer.

In pulling tests with the finger the three PDMS mixture ratios are tested. For mixture ratio 1:2 and 50:1, all test are unsuccessful (fig. 3.5). however, at mixture ratio 4:1, a successful detachment was performed. However, the detachment cannot be confirmed as the before picture did not cover the area of the missing piece (fig. 3.11). without this proof, the formation of the hole when freezing cannot ruled out even if unlikely. Also a piece of ice is not distinguishable when hanging on the finger. Still, since a mixture ratio gave partial succesful results, this mixture ratio needs to be investigated further.

To test the behavior of an ice layer frozen onto PDMS, the surface is damaged in two ways. In one experiment, the Sample is scratched with a diamond pencil. This is done after mounting the sample on the shuttle. The diamond tip is drawn around the inside of the window several times, forming a visible groove (fig. 3.12a). In the other experiment, The pincer is intentionally placed through the middle of the sample when plunge-freezing. Therefore an intentional big imprint is left after freezing (fig. 3.12b).


Scratching the ice layer with a diamond pencil

Category	Hand-freezed	Plunge-freezed
count executed tries	4	4
unsuccessful	3	3
breaks/movement of ice	1	1
piece lifted with finger	0	0


Table 3.4: Comparison of detachability between hand-freezed and plunge-freezed samples

Category	PDMS 1:2	PDMS 4:1	PDMS 50:1
count executed tries	5	3	6
unsuccessful	5	2	6
breaks/movement of ice	0	0	0
piece lifted with finger	0	1*	0


Table 3.5: Comparison of detachability between PDMS mixture ratios. Pulling tests at PDMS of 1:2 and 50:1 mixture ratio were unsuccessful. But at mixture ratio of 4:1, detachment was successful one time. However, this could be a statistical outlier or a false positive. A hole was spotted after pulling with the finger. however, the hole is located outside of the before image. therefore the hole could already exist while freezing even if unlikely.



(a) PDMS 1:2 sample in true light




(b) PDMS 1:2 sample with fluorescence filter




(c) PDMS 4:1 sample in true light



(d) PDMS 4:1 sample with fluorescence filter



(e) PDMS 50:1 sample in true light



(f) PDMS 50:1 sample with fluorescence filter

Figure 3.10: comparison of ice on different PDMS samples. The ice crystallized ice structure only changes minimally. A continuous ice layer forms over each plasma activated PDMS sample. the layer is only interrupted by the imprint of the pinces at plunge-freezing here visible in (f).




Figure 3.11: Successful pulling test of 4:1 PDMS. However, this result should be taken cautiously. The before and after image each catch a different part of the sample. Under normal circumstances, no holes form in the ice layer when freezing (shown in fig. 3.10) however, this case cannot be ruled out.




Figure 3.12: Twod different methods to break the continuos ice surface.

3.2.7 observed error sources

(AAH ABKLÄREN MIT SECTION IN METHODE: VIELLEICHT ALLES HIER REIN ODER VORNE)

When attaching, the right distance between sample and "finger" helps increasing the tensile force which can be applied to the sample. As discussed in section 2.1.4, two types of positioning errors can occur: The positioning over the sample and the gap between sample and "finger". The first error can be corrected by stages in X and Y directions. The other error also depends on the HFE volume. With little volume, there is no risk of HFE spreading between sample and "window". With a high volume, decreasing the gap can result in HFE spreading under the window brace if the "finger" is lowered too much. On the other hand, leaving a big gap between "finger" and sample results in a thicker HFE layer which takes less tensile forces.

The duration of time between reaching the "glue" state from the "unglue" state makes disconnecting before pulling more likely. The sample is cooled with the "finger" on top as well as by the harbor cooled with liquid nitrogen. Best case is the same cooling rate of "finger" and harbor. In practice, the cooling of the "finger" is faster. This also leads to successful attaching. If for example major leaks occur in the cold nitrogen gas supply, the cooling rate of the "finger" is slower than the harbor. In those cases, the "finger" does not stay attached to the sample. A theory is that HFE gets more viscous on the sample than on the "finger". While the HFE is perfectly adhering to the samples surface. Additionally, the HFE loses volume due to cooling. The HFE on the tip of the "finger" is able to fill out the missing volume and thin out around the finger due to gravity. With enough time, this could be the HFE enough that the finger is not properly attached anymore. (NICHT SO DIE BESTE THEORIE; EIGENTLICH BEI GRO?ER LÜCKE?)

The formation of ice inside the bath is largely inhibited through the Nitrogen Gas inlets. The nitrogen is forming a barrier to the atmosphere which contains humidity that turns into ice at cold temperatures. However, through turbulence and the finger, some ice can form inside the bath.

In some experiments, high buildup of additional ice is observed on the sample. The additional ice is looser, reducing the possible grip onto the desired ice layer. The formation is traced back to a leakage in the finger. The tolerances between 3D printed part and metal bar should be able to seal the gap airtight. But through changing out parts, the tolerances are looser, allowing a weak cold nitrogen current right on the sample. There are two possibilities where the humidity itself is coming from: The cold nitrogen gas itself could contain some humidity. Second through additional turbulence, more air is sucked through the barrier. The air is directed directly on the sample, causing more ice buildup. To avoid this, the gap is sealed tight with twinsil lubricating silicone typically used for dental application.

3.3 Inverted microscope

The used microscope setup is a prototype based on an uninverted microscope. As using more inverted light microscopes are considered for usage in cryogenic temperature, the setup is analyzed and some common issues and possible fixes are proposed.

To take pictures, a two-fold and 5-fold objective is used. The 5-fold objective fits between the microscope and glasses assembly without a significant gap. However, to fit the 2-fold objective the glasses assembly needs to be designed thinner. With the increased gap between objective and glasses assembly, more ambient light is lighting the sample. This decreased the visibility of fluorescence on samples. With the increased field of view with the 2-fold objective, total reflection of the two glass panes in the glasses assembly is inducing strong noise in the pictures. Also the access hole to the harbor is a bit too small. To reduce noise and increase the

harbor access, the glasses assembly is fixed tilted to the box by adding spacers. In future, the glasses assembly needs to be redesigned to fix all those issues.

In general, fitting the sample into the harbor is difficult. The cold gas is exhausted over the access to the harbor. The exhausted cold gas results in fog and locally reduces oxygen. The big stage and cooled box on top reduces visibility to the access point. In future, the exhaust should be pointed away from important access points.

Ice frozen to the assembly can also damage the microscope. Especially when the setup warms back up to room temperature, the ice melts and water drops onto important parts. Therefore, the objectives should be designed water proof. Gaps should be closed as much as possible to avoid water entering the inside of the microscope.

4 Conclusion and Outlook

The finger with HFE is able to apply tensile stress between 103.3 kPa and 585.4 kPa. The stress applied is a couple magnitude higher than the stress calculated for detaching PDMS. The discrepancy could be result of energy lost deforming HFE at cryogenic temperatures and the stability of a continuous ice layer.

Different factors which increase and decrease the effectiveness of the finger: Too much HFE results in HFE ripping because of the low cohesion. Too little HFE is prone to not properly attaching to the Sample. lower temperature increase the viscosity and therefore stability of HFE. But at around -170°C , cracks form in HFE and the strength is drastically reduced. The direction of force could help at detaching an ice layer. However, the observed stability to shear load of HFE is considerably lower than to tensile load. The ice structure is not negletable. A continuous ice layer has more adhesion to the sample than an already broken layer. Positioning is also important, especially keeping the correct gap between sample and finger in which the HFE is spread thin but not flows between window and sample. Also slow temperature changes are drastically reducing the likelihood of a good connection between sample and finger.

At room temperature, the PDMS with different mixture ratios were tested. A mixture ratio of 1:2 has shown low adhesion forces as predicted. However, the transfer of results at room temperature to ice adhesion at cryogenic temperatures is only possible with restrictions. Plasma activation increases the adhesion forces potentially more on ice than the UV glue. At other mixture ratios, plasma treatment leads to brittle surfaces, which is not observed for 1:2 mixture ratio. Attempts of detaching an ice layer off this PDMS coated sample were not successful.

Additionally, A PDMS mixture of 4:1 and 1:50 were investigated. With plasma treatment with 100 % power and 10 min, Cracks form and the surface is brittle. However, The ice layer frozen on top of a PDMS coated sample is continuous. Ions produced in plasma treatment increase the adhesion force too much for the finger to detach. In future, stress needed to break the ice layer should be taken into account. Also to decrease the area of contact to the slide, a grid could be used to reverse the hydrophilic effect of plasma activation before freezing. The resulting small pieces of ice may be easier to detach than a clamped down continuous layer.

In contrast to PDMS coated slides, lipid coated slides resulted in cracks in the ice layer itself. In pulling tests with the finger, breaking and moving parts of the ice layer were possible in 1 out of 4 cases. To help detach the lipid layer, PDMS could be used. With increased brittleness and pre broken ice layer, less stress is needed for detaching.

In Experiments, no lipid solvent combination was found with high solubility at cryogenic temperatures. All tested solving processes are endothermic. In general, endothermic processes are inefficient in cold temperatures.

The potential for lipids and detergents is not exhausted in this thesis. lipids and detergents can be engineered for lower adhesion forces. Additionally, finding solvents by solely experiments is very unlikely.

To engineer a sacrificial layer, other detergents could be used [21]. An exothermic process is also not limited through the cyrogenic temperatures. Alternatively, to increase solubility, changing the pH value of lipid and solvent could increase solubility [22].

The force applied of the "finger" has proven to be not enough to break any surface on the sample. Pre broken pieces are sometimes picked up. A process is needed to completely loosen the ice layer without the "finger". For example, deactivating the plasma by putting a grid on the PDMS after Plasma activation could result in a loose, incontinuous but regular layer. Smaller pieces with less adhesion are easier to detach.

In general, the finger setup is not reliable. Even after analysis, forces applied with the finger vary too much for inducing strong forces to a sample. However, pre broken pieces may be able to be picked up easier. Therefore, other methods of breaking loose the ice layer should be searched.

Another way to create a loose ice layer is to engineer the ice layer itself. Additives inside the ice layer could help in the breaking process. Freezing the ice in an emulsion could result in an incontinuous ice layer.

5 References

- [1] Rainer Heintzmann and Gabriella Ficz. “Breaking the resolution limit in light microscopy”. In: *Briefings in functional genomics & proteomics* 5.4 (2006), pp. 289–301. issn: 1473-9550. doi: 10.1093/bfgp/e11036.
- [2] Dganit Danino. “Cryo-TEM of soft molecular assemblies”. In: *Current Opinion in Colloid & Interface Science* 17.6 (2012), pp. 316–329. issn: 1359-0294. doi: 10.1016/j.cocis.2012.10.003. url: <https://www.sciencedirect.com/science/article/pii/S1359029412001033>.
- [3] R. Faoro, M. Bassu, and T. P. Burg. “Determination of the refractive index of liquids at cryogenic temperature”. In: *Applied Physics Letters* 113.8 (2018), p. 081903. issn: 0003-6951. doi: 10.1063/1.5043370.
- [4] Brian Wowk. “Thermodynamic aspects of vitrification”. In: *Cryobiology* 60.1 (2010), pp. 11–22. issn: 0011-2240. doi: 10.1016/j.cryobiol.2009.05.007. url: <https://www.sciencedirect.com/science/article/pii/S0011224009000674>.
- [5] Siyan Yang et al. “Condensation frosting and passive anti-frosting”. In: *Cell Reports Physical Science* 2.7 (2021), p. 100474. issn: 2666-3864. doi: 10.1016/j.xcpr.2021.100474. url: <https://www.sciencedirect.com/science/article/pii/S2666386421001740>.
- [6] Marc P. Wolf, Georgette B. Salieb-Beugelaar, and Patrick Hunziker. “PDMS with designer functionalities—Properties, modifications strategies, and applications”. In: *Progress in Polymer Science* 83 (2018), pp. 97–134. issn: 0079-6700. doi: 10.1016/j.progpolymsci.2018.06.001. url: <https://www.sciencedirect.com/science/article/pii/S0079670017300783>.
- [7] Junpeng Liu et al. “Development and evaluation of poly(dimethylsiloxane) based composite coatings for icephobic applications”. In: *Surface and Coatings Technology* 349 (2018), pp. 980–985. issn: 02578972. doi: 10.1016/j.surfcoat.2018.06.066.
- [8] Lasse Makkonen. “Ice Adhesion —Theory, Measurements and Countermeasures”. In: *Journal of Adhesion Science and Technology* 26.4-5 (2012), pp. 413–445. issn: 0169-4243. doi: 10.1163/016942411X574583. url: https://www.researchgate.net/publication/254214058_Ice_Adhesion_-Theory_Measurements_and_Countermeasures.
- [9] Raffaele Faoro et al. “Aberration-corrected cryoimmersion light microscopy”. In: *Proceedings of the National Academy of Sciences of the United States of America* 115.6 (2018), pp. 1204–1209. doi: 10.1073/pnas.1717282115.
- [10] DOW. *Datasheet SYLGARD™ 184 Silicone Elastomer*. Ed. by DOW.
- [11] Guiguan Zhang et al. “Experimental study on mechanical performance of polydimethylsiloxane (PDMS) at various temperatures”. In: *Polymer Testing* 90 (2020), p. 106670. issn: 01429418. doi: 10.1016/j.polymertesting.2020.106670.
- [12] Alexandra Borók, Kristóf Laboda, and Attila Bonyár. “PDMS Bonding Technologies for Microfluidic Applications: A Review”. In: *Biosensors* 11.8 (2021). doi: 10.3390/bios11080292.

- [13] Michael J. Owen and Patrick J. Smith. "Plasma treatment of polydimethylsiloxane". In: *Journal of Adhesion Science and Technology* 8.10 (1994), pp. 1063–1075. ISSN: 0169-4243. doi: 10.1163/156856194X00942.
- [14] Taiki Ohishi et al. "Tensile strength of oxygen plasma-created surface layer of PDMS". In: *Journal of Micromechanics and Microengineering* 27.1 (2017), p. 015015. ISSN: 0960-1317. doi: 10.1088/0960-1317/27/1/015015.
- [15] Pablo F. Ibáñez-Ibáñez et al. "Ice adhesion of PDMS surfaces with balanced elastic and water-repellent properties". In: *Journal of Colloid and Interface Science* 608.Pt 1 (2022), pp. 792–799. ISSN: 0021-9797. doi: 10.1016/j.jcis.2021.10.005. URL: <https://www.sciencedirect.com/science/article/pii/S0021979721016714>.
- [16] Srirama M. Bhairi, Ph. D. "Detergents: A guide to the properties and uses of detergents in biology and biochemistry". In: (2001), pp. 6–7.
- [17] Mark Tye Zafir Javeed. *Enthalpy of Solution*. URL: [https://chem.libretexts.org/Bookshelves/Physical_and_Theoretical_Chemistry_Textbook_Maps/Supplemental_Modules_\(Physical_and_Theoretical_Chemistry\)/Physical_Properties_of_Matter/Solutions_and_Mixtures/Solution_Basics/Enthalpy_of_Solution](https://chem.libretexts.org/Bookshelves/Physical_and_Theoretical_Chemistry_Textbook_Maps/Supplemental_Modules_(Physical_and_Theoretical_Chemistry)/Physical_Properties_of_Matter/Solutions_and_Mixtures/Solution_Basics/Enthalpy_of_Solution).
- [18] Charles E. Mortimer and Ulrich Müller. *Chemie: Das Basiswissen der Chemie ; 126 Tabellen*. 9., überarb. Aufl. Stuttgart: Thieme, 2007. ISBN: 9783134843095. URL: https://books.google.de/books?id=kf_1FH8Le6QC.
- [19] R Wayne Albers. "Phospholipid Bilayers". In: *Basic Neurochemistry: Molecular, Cellular and Medical Aspects. 6th edition*. Ed. by R. Wayne Albers. Lippincott-Raven, 1999. URL: <https://www.ncbi.nlm.nih.gov/books/NBK28248/>.
- [20] PubChem. *Ethane*. 2023. URL: <https://pubchem.ncbi.nlm.nih.gov/compound/Ethane> (visited on 08/29/2023).
- [21] Sigma-Aldrich. *Safety data sheet HFE 7200*. Ed. by Sigma-Aldrich. 2023.
- [22] Bruce A. Averill, Patricia Eldredge. *Principles of General Chemistry: Solubility and pH*. URL: <https://2012books.lardbucket.org/books/principles-of-general-chemistry-v1.0/s21-04-solubility-and-ph.html>.