Tests of measurements in vacuum, SAF P10695

Test measurements, to see if we can measure at RT in the cryo

28/04/2022

Sample P10695 at RT
Tip FR076-3N7
Léa and Aurore measuring

- Issues with the setup, max PL 50 kcts/s after changing the fiber for an old one.
- Wire attached on the sample.
- ODMR scan 1016-34: MW power 18 dBm, laser power lambda/2 at 60° (PL 35 kcts/s), retracted :



- ODMR scan 1016-34: MW power 25 dBm, laser power lambda/2 at 60° (PL 35 kcts/s), retracted :



- ODMR scan 1016-34: MW power 24 dBm, laser power lambda/2 at 75° (PL 40 kcts/s), retracted :



- ODMR scan 1016-34: MW power 24 dBm, laser power lambda/2 at 55° (PL 24 kcts/s), retracted :



- ODMR scan 1016-34: MW power 24 dBm, laser power lambda/2 at 65° (PL 24 kcts/s), retracted :



- ODMR scan 1016-34: MW power 24 dBm, laser power lambda/2 at 45° (PL 14 kcts/s), retracted :



- ODMR scan 1016-34: MW power 24 dBm, laser power lambda/2 at 60° (PL 30 kcts/s), retracted :



- ODMR scan 1016-34: MW power 22 dBm, laser power lambda/2 at 60° (PL 30 kcts/s), retracted :



- ODMR scan 1016-34: MW power 22 dBm, laser power lambda/2 at 75° (PL 40 kcts/s), retracted :



- ODMR scan 1016-34: MW power 22 dBm, laser power lambda/2 at 50° (PL 15 kcts/s), retracted :



- ODMR scan 1016-34: MW power 20 dBm, laser power lambda/2 at 50° (PL 15 kcts/s), retracted :



- ODMR scan 1016-34: MW power 20 dBm, laser power lambda/2 at 45° (PL 10 kcts/s), retracted :



- ODMR scan 1016-34: MW power 20 dBm, laser power lambda/2 at 55° (PL 15 kcts/s), retracted :



- ODMR scan 1016-34: MW power 18 dBm, laser power lambda/2 at 50° (PL 15 kcts/s), retracted :



- ODMR scan 1016-34: MW power 16 dBm, laser power lambda/2 at 60° (PL 30 kcts/s), retracted :



- ODMR scan 1016-34: MW power 16 dBm, laser power lambda/2 at 55° (PL 25 kcts/s), retracted :



- ODMR scan 1016-34: MW power 16 dBm, laser power lambda/2 at 80° (PL 42 kcts/s), retracted :



- ODMR scan 1016-34: MW power 25 dBm, laser power lambda/2 at 70° (PL 38 kcts/s), retracted :



- ODMR scan 1016-34: MW power 25 dBm, laser power lambda/2 at 50° (PL 15 kcts/s), retracted :



- ODMR scan 1016-34: MW power 25 dBm, laser power lambda/2 at 55° (PL 20 kcts/s), retracted :



- ODMR scan 1016-34: MW power 25 dBm, laser power lambda/2 at 65° (PL 28 kcts/s), retracted :



- TEST : ODMR scan 1016-34: MW power 25 dBm, laser power lambda/2 at 60° (PL 30 kcts/s), retracted :



*Pause d'un aimant :