TECHNICAL SPECIFICATIONS

NUCLEAR DATA INC
100 West Golf Road
Palatine, Illinois 60067
Phone A.C. 312 • 529-4600



SERIES 3300 SYSTEM POWER SUPPLY

The Series 3300 System Power Supply provides all necessary voltages for operating the Series 3300 System analyzer. Features include a meter for monitoring all voltages and indicator lamps for overload conditions, high temperature conditions and A.C. line. Automatic shutoff is provided when overload or high temperature conditions are encountered. All voltages are screwdriver adjustable from the rear panel.

Power Requirements

Approximately 760 watts.

Line Voltage Requirements

115 volts A.C. ±10%, 60 Hz.
220 volts A.C. (available upon request), 50 or 60 Hz.

Outputs

+6 volts D.C. at 12 amperes.
−6 volts D.C. at 6 amperes.
+15 volts D.C. at 2 amperes.
−15 volts D.C. at 2 amperes.
+22 volts D.C. at 2.5 amperes.
−22 volts D.C. at 2.5 amperes.
+48 volts D.C. at 1.5 amperes.
−48 volts D.C. at 0.5 amperes.

Ripple

Less than 10 millivolts peak-to-peak on all outputs.

Temperature Range

The power supply will operate satisfactorily within the range of 0° to 50° Centigrade.

Regulation

Output	Load	Line	Output	Load	Line
+6 VDC	0.5%	0.25%	+22 VDC	0.1%	0.1%
−6 VDC	1.0%	0.5%	−22 VDC	0.1%	0.1%
+15 VDC	0.2%	0.1%	+48 VDC	0.1%	0.05%
−15 VDC	0.2%	0.1%	−48 VDC	0.1%	0.05%

Fusing

The line input is fused with two 10-ampere slow blow fuses. All outputs are electronically fused.

Weight

Approximately 67 pounds.

Dimensions

Approximately 17" wide x 7" high x 21" deep.

Part Number

74-0064

SPECIFICATIONS SUBJECT TO CHANGE WITHOUT NOTICE
PRINTED IN U.S.A. 11/67