# CERTIFICATE OF CALIBRATION

**Certificate Number :** **EXPR-THL-493-01** **Job Order Number : 493**

**Customer Details :**

**Sinohydro Corporation Limited Branch**
**Al Muhammadiyah Dist,Kingdom of Saudi Arabia**

**Instrument Details of UUC :**

| | | |
|---|---|---|
| Description | : | Infrared Thermometer |
| Manufacturer | : | Raythink |
| Manufacturer Model Number | : | CX Series |
| Manufacturer Serial Number | : | CB4360250 |
| ID. Number | : | NA |
| Range | : | -20 to 550 °C |
| Resolution/Least Count | : | 0.1 °C |
| Accuracy | : | ±2 °C or ±2% of rdg whichever greater |
| Visual Inspection | : | ok |
| Location of the Device | : | NA |
| Calibration Location | : | Expert Power Temperature Lab |
| Received Date | : | 27-Jun-26 |
| Certificate Issue Date | : | 4-Jul-26 |
| **Calibration Date** | : | **2-Jul-26** |
| **Recommended Due Date** | : | **2-Jul-27** |

**Instrument Conditions of UUC :**

**As Found:** Within Tolerance **As Left:** Calibrated within Tolerance

**Environmental Conditions at the Calibration Location :** Expert Power Temperature Lab

| | | |
|---|---|---|
| Ambient Temperature | : | 23 ± 2 °C |
| Relative Humidity | : | 50 ± 10 % RH |
| **Calibration Method:** | : | EXP-CP-TH-04 |

**Details of the Standard Equipment Used for Calibration :**

| Description | Sl. No / ID. No | Certificate No | Validity |
|---|---|---|---|
| IR Calibrator | I-IR20025 | FL/C/TH/28022026-C019 | 28-Mar-27 |
| IR Calibrator | K-IR5058 | FL/C/TH/24022026-C066 | 25-Feb-27 |

**Note:**

1. The Certificate refers only to the particular item(s) submitted for calibration.
2. All Results Reported are valid at the time of and under the stated conditions of measurements.
3. The equipment/tool has been calibrated using measurement standards with traceability of measurement to national or international standards, which realize the International system of Units(SI)
4. This certificate shall not be reproduced, without the written permission of Expert Power Calibration Laboratory.
5. The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor (k=2) such that the coverage probability corresponds to approximately 95%. The uncertainty evaluation has been carried out in accordance with JCGM 100:2008 "Guide to the Expression of uncertainty in Measurement".
6. Expert Power Calibration Laboratory (An ISO 9001:2015 QMS certified company) hereby certifies that this product was calibrated/verified using procedures developed by Expert Power Calibration Lab in accordance with the specifications of the manufacturer and conform to ISO 17025:2017.

**Calibration Technician**

EP 110 CAL LAB

**Calibration Tech. Stamp**
**RPT-TCC-003v2.0**

CALIBRATED EXPERT POWER Calibration and Maintenance

**Authorised Signatory**

**Shareef Yousaf**
**( Technical Manager)**



Expert Power for Trading Est.
Sitteen Street, Malaz P.O. Box, 32933,
Riyadh, 11438, Saudi Arabia
Tel-Fax: +966 11 421 0481

مؤسسة خبير الطاقة للتجارة
شارع الشتين الملزص.ب ٣٢٩٣٣ الرياض ١١٤٣٨
المملكة العربية السعودية تلفون فاكس
+٩٦٦ ١١ ٤٢١ ٠٤٨١

# Calibration Results

**Certificate Number :** EXPR-THL-493-01 **Job Order Number :** 493

| Calibration Point | Reference Value | | Average of UUC Reading | | Error | | Expanded Measurement Uncertainty(±) | |
|---|---|---|---|---|---|---|---|---|
| -20 °C | -20.1 | °C | -19.3 | °C | 0.8 | °C | 2.0 | °C |
| -10 °C | -10.1 | °C | -10.4 | °C | -0.3 | °C | 2.0 | °C |
| 0 °C | 0.2 | °C | 0.1 | °C | -0.1 | °C | 2.0 | °C |
| 50 °C | 48.3 | °C | 48.2 | °C | -0.1 | °C | 1.7 | °C |
| 100 °C | 98.9 | °C | 98.4 | °C | -0.5 | °C | 1.8 | °C |
| 200 °C | 197.6 | °C | 196.0 | °C | -1.6 | °C | 1.9 | °C |
| 300 °C | 297.8 | °C | 296.0 | °C | -1.8 | °C | 1.9 | °C |
| 400 °C | 396.2 | °C | 394.0 | °C | -2.2 | °C | 2.0 | °C |
| 500 °C | 495.7 | °C | 492.8 | °C | -2.9 | °C | 2.0 | °C |

**Conclusion / Remarks:**

1. Measurement Uncertainty at 95.45 % Confidence level and Coverage Factor k=2.
2. The above Results are based on either Customer Request or manufacturer with in specified accuracy limits.
3. Calibration points marked "*" are not within the scope of Accreditation.
4. The calibration result is within the tolerance limits (if listed) based on simple acceptance in which uncertainty is not considered. Unless otherwise stated calibration data on certificates is Calibration Result . It is the responsibility of the end user to determine if it is appropriate for their specific application. NO statement of conformity is stated in this report

E.P 110 CAL LAB

**End of Certificate**



Expert Power for Trading Est.
Sitteen Street, Malaz P.O. Box, 32933,
Riyadh, 11438, Saudi Arabia
Tel-Fax: +966 11 421 0481

مؤسسة خبير الطاقة للتجارة
شارع الشتين الملزص.ب ٣٢٩٣٣ الرياض ١١٤٣٨
المملكة العربية السعودية تلفون فاكس
+٩٦٦ ١١ ٤٢١ ٠٤٨١