MIM Maestro 7.4.3 — Contouring Quick Reference

Hotkeys, custom shortcuts & tools · with GenesisCare GCC-GDE-013

Dr Lincoln Dinh · November 2026

About this sheet

- Default shortcuts, menu paths and tool behaviour come from the GenesisCare guide **GCC-GDE-013 v1.1** (*MIM – Introduction and Functionality Guide*, last reviewed Sept 2022). It predates 7.4.3, so check a default in **Settings › Keyboard Shortcuts…** if it doesn't behave as listed.

1 · Default keyboard shortcuts

Key	Display & navigation
`F1`	W/L preset: **Soft tissue**
`F2`	W/L preset: **Lung**
`F3`	W/L preset: **Bone**
`F4`	W/L preset: **Brain**
`W`	Manual window / level for any dataset
`3` / `4`	Zoom in / zoom out
`=`	Toggle crosshair display
`↑` `↓`	Scroll image slices
`←` `→`	Move between session pages

Key	Editing & tools
`Ctrl` + `Z`	Undo (back to the last saved point)
`Ctrl` + `Y`	Redo
`C`	**Clean** current slice of the contour
`X`	**Erase** current slice of the contour
`F`	**Fill** current slice of the contour
`E`	Localise to a DICOM coordinate
`M`	MIP movie / 3D render — check bolus covers wires, final OAR review
`Ctrl` + `Shift` + `B`	Transfer contours to another dataset (fused or same frame of reference)

2 · Mouse & modifier actions

Action	What it does
Brush: **left-click drag**	Draw or erase contours (2D and 3D brush)
Brush: **right-click drag**	Resize the brush
`Alt` held while brushing	Draw **outside** the current contour — creates a separate island (e.g. bilateral nodes) instead of merging
Dynamic brush	Left-click drag, keeping the **inner circle inside the organ of interest**
Pan tool	Right-click drag to move the image
Zoom tool	Left-click in the image, drag up / down
Double-click contour name	Show contour statistics (volume, density); double-click again to hide
Right-click contour name	Set as **Hard** contour (H appears); right-click the H to remove it
Double-click an image	Creates a new page in the session from that view
`Shift` + click (patient list)	Combine search results for the same patient under different IDs (e.g. PET from another site)

3 · Create your own hotkeys

Settings › Keyboard Shortcuts… › Edit

→ Click **Settings** (cog, top-left bar next to Patients / Sessions).
→ Choose **Keyboard Shortcuts…** — the list shows every action and its current key.
→ Select the action, click **Edit**, press the new key or combination, then confirm.
→ Avoid keys already taken: F1–F4, W, 3, 4, =, C, X, F, E, M, arrows, Ctrl+Z / Y, Ctrl+Shift+B.
→ Your settings are stored per user on the MIM server (network-user settings).

Worth binding (no default key in the guide)

- 2D brush · 3D brush · Dynamic brush
- Interpolate · Co-Pilot
- Smooth · Hard Edge · Clone
- Toggle adjacent contour slices
- Expansion / Contraction · Boolean
- Your most-used site workflow (e.g. *GC SBRT Lung – Targets*)

Steps 3–5 describe the dialog as it usually works; the guide gives only the menu path, so wording may differ slightly in 7.4.3.

4 · Set up your workspace

What	How	Tip
Toolbar	Cog on the right of the toolbar → green **+** add, red **–** remove, drag to reorder Tools, Extensions and Workflows. Search bar finds functions by keyword. Saved for next login.	Reset to site default: MIM SME / physics deletes your user settings file on the MIM server.
Mode Tools	Settings option in the Mode Tools panel → add / remove contouring tools or workflows.	Keep only the 6–8 tools you use.
Contour settings	Colour, line width, name, **Fill** 0–100 % (good for checking Hard Edge results), **Range lock** with bone / tissue / air presets or custom HU.	Range lock: set back to *None* after use.
Contour cog menu	Rename, erase, delete, clone, statistics, **Localize to Contour Centroid**, **Lock Contour**.	Lock previous-treatment volumes so they can't be edited by mistake.
Series to Contour	Must always be the **Planning CT** — structures attach to this dataset.	Check before you start drawing.
Colour / W-L	Dataset cog → contrast presets; click the grey-scale bar to change colour map (fusions, PET).	`W` for custom W/L.

5 · Contouring toolkit

Tool	How it works	Use it for
2D / 3D brush	2D: one slice, any plane. 3D: across slices. Hold Alt to start a separate contour.	Most targets and OARs; 3D brush for globes, mandible.
Dynamic brush	Left-click drag; keep the inner circle inside the organ of interest.	Organs with a clear boundary.
Co-Pilot	Scroll to the next slice, wait for the shadowed suggestion, left-click to accept, edit, repeat. Uses deformable registration.	Parotids, cord, brainstem, oesophagus.
Interpolate	Contour every few slices, click to fill the gaps; can be re-run on the same structure.	Tubular OARs. Tip: *Erase Contour Slices* (every other slice) then interpolate to tidy a jagged structure.
Toggle adjacent slices	Shows the current contour dotted on the previous / next slice.	Consistency slice to slice; 3D bolus.
Clean & Fill	Clean removes small islands. Fill: single slice, each slice, or volumetric (only fully enclosed holes).	After range-lock contouring: range lock None → Fill → Smooth.
Smooth	One click smooths edges; repeatable.	Over-smoothing shrinks the volume.
Hard Edge	Removes overlap with any contour marked Hard (H).	Bowel bag minus bladder; pharynx vs larynx / constrictors.
Expand / Contract	Uniform or six-direction margins; inner / outer rings.	PTVs, PRVs, skin and bolus rings.
Boolean	Contour → operation → contour (chain several) → destination → green tick; steps can be undone.	PTV minus OAR, combined lungs.
Region Grow	Load Region Templates (e.g. Chest › Rt Lung) → adjust box → chequered flag → Done.	Lungs — then delete trachea / contralateral air.
Whole Body · Clone · Erase	Body contour by dragging range in sag / cor · duplicate contour · erase all slices (name kept, in italics).	Erase ≠ delete: name stays in the list.

6 · Head & neck OAR recipes (GenesisCare tips)

Structure	Recipe
Parotids	Move tool, or 2D brush + Co-Pilot; range lock *tissue*
Submandibular glands	Move tool + 2D brush with Co-Pilot or Interpolate
Globes / Lens	Globes: Move tool + Smooth, or 3D brush. Lens: 2D brush + Smooth
Brainstem · Spinal cord · Oesophagus	2D brush + Interpolate or Co-Pilot. Consider *Erase Contour Slices* (every other) then Interpolate
Larynx · Pharyngeal constrictors	2D brush + Interpolate, then Smooth; set as Hard Edge to avoid overlap
Pharynx	2D brush + Interpolate or Co-Pilot, Smooth; avoid overlap with larynx, constrictors and oral cavity
Mandible	Range lock *bone* → 2D or 3D brush → range lock *None* → Fill & Smooth
Optic nerves	2D brush + Smooth; consider custom range lock lower −50 HU

7 · Sessions, saving & recovery

Saving

- Blue floppy in the contour list saves **structures** to a dataset — it does **not** save the session.
- **NSW:** save sessions with the *Save Session MIM Workflow* (local work instruction).
- Contour names in *italics* are empty; white names contain contours.
- Avoid several open sessions for one patient — it causes confusion.
- Use **Session Notes** to flag issues for RTs / planners.

Recovery & auto-deletion

- **Autosave recovery:** Import › Browse… → your state's MIM recovery share (ask your MIM SME) → your username › RECOVERY › RTAUTOSAVE → select all RTst files → MIMpacs → reopen with *Add All Data*.
- Data are **auto-deleted after 35 days** without opening (12:01 am daily).
- To keep a case: save a session called **Keep**, or anonymise with **Case** in the patient name (teaching).

8 · My custom hotkeys

Key	Action	Notes

Sources: GenesisCare GCC-GDE-013 *MIM – Introduction and Functionality Guide* v1.1 (Oncology Australia), incl. *MIM Tips & Tools – Head & Neck*; MIM Software product information (Co-Pilot). Personal reference — follow current GenesisCare work instructions where they differ.