Illuminating Neural Dynamics:

The Physics of Viral-Mediated Two-Photon Calcium Imaging

Qingyang Ling, Ian Dong, Finlay McCormack

School of Physics and Astronomy, The University of Edinburgh

Abstract

Modern neuroscience requires observing circuit-level activity across large neuronal populations in intact tissue. Viral-mediated two-photon calcium imaging has emerged as a transformative solution to achieve this, overcoming historical limitations of targeted gene delivery and tissue scattering. In this review, we synthesize the complete physical pipeline of this state-of-the-art technique, demonstrating how the interplay of viral delivery, molecular photophysics, and non-linear optics ultimately enables the successful, high-resolution observation of neural population dynamics in vivo.

INTRODUCTION

Extracting dynamic neural signals from the intact brain is fundamentally a complex biophysical and physical problem. To systematically explore the mechanisms underlying viral-mediated calcium imaging, we deconstruct the entire measurement pipeline into four sequential sections. Section 1 examines the targeted delivery of genetic cargo, modeling the brain's extracellular space to quantify steric and electrostatic constraints. Section 2 evaluates the GECI transducer, coupling the thermodynamics of calcium binding to the photophysics of fluorescence. Section 3 addresses signal readout, detailing how two-photon microscopy leverages non-linear absorption to mitigate highly scattering brain tissue. Finally, Section 4 explores how dimensionality reduction techniques decode low-dimensional neural manifolds from these physically filtered traces.

```
Action Potential
        ↓
Voltage-gated Ca²⁺ channels open
        ↓
Ca²⁺ influx
        ↓
GCaMP conformational change
        ↓
Fluorescence increase
        ↓
Two-photon detection
```

Figure 1: The signal transduction cascade in GECI-based calcium imaging.

1 VECTOR DELIVERY UNDER BIOLOGICAL & PHYSICAL CONSTRAINTS

1.1 Biological Principles: Gene Delivery to Illuminate Neuronal Activities

Neurons communicate through fast electrical events called action potentials. During these events, membrane depolarization causes voltage-gated calcium channels to open, driving a rapid influx of Ca^{2+} into the cytosol. To convert these calcium transients into measurable fluorescence changes, an intracellular transducer is required. This is exactly what Genetically Encoded Calcium Indicators (GECIs), such as the GCaMP family, are designed to do. GCaMP is a protein sensor whose fluorescence depends on calcium binding: when intracellular Ca^{2+} rises, binding induces a conformational change that significantly increases fluorescence intensity [1] . With these fluorescence transients available, two-photon microscopy calcium imaging becomes a practical platform for in vivo readout, as it can record from deep within scattering brain tissue while maintaining single-cell resolution over a wide field of view [2].

Crucially, GCaMP is a protein and must be synthesized inside target neurons rather than delivered as a freely diffusing molecule. Achieving expression therefore requires delivering the genetic construct which encodes GCaMP into target cells. However, naked DNA is large, highly negatively charged, and readily degraded, meaning it cannot efficiently cross biological barriers on its own. Consequently, success-

ful GECI-based calcium imaging depends fundamentally on an effective vector strategy that protects the DNA payload, transports it through brain tissue, and enables expression in the targeted neuronal population [5].

1.2 Physical Barriers to Delivery: Brain Tissue as a Porous Medium

The immediate physical bottleneck for any viral vector is extracellular transport. Neurons and supporting glia cells are not suspended in “free water”; instead, molecules must move through the brain extracellular space (ECS), a crowded and structured interstitial network well captured by a porous-medium description [3].

To understand the severity of this barrier, we can visualize the ECS through two governing physical parameters. First, the volume fraction is typically $\alpha \approx 0.2$, meaning that cellular structures occupy roughly 80% of the tissue volume, leaving only a scant 20% as narrow, interconnected channels for transport. Second, the tortuosity is typically $\lambda \approx 1.6$ for small diffusing species. Rather than purely quantifying geometric complexity, λ is an empirical hindrance factor that summarizes how the ECS microstructure slows diffusion: vectors are forced to circumnavigate cellular obstacles, following winding, tortuous paths rather than free paths in free solutions [3].

Crucially, the effective width of these extracellular channels is estimated to be only 38–64 nm [4], implying a strongly size-dependent hindrance. Furthermore, the extracellular matrix (ECM) is not chemically neutral: it carries abundant fixed negative charges and provides reversible interaction sites. Ligands with affinity to ECM components, such as sugar polymers like heparan sulfate, experience a reactive "stickiness" that can substantially slow effective transport [8] [9].

As a result, transport in brain tissue is better viewed as navigation through an evolving, structured, and chemically active maze [7]. This cumulative hindrance is captured by the relationship: $D^* = D/\lambda^2$ [3]. Through a combination of physical trapping and electrostatic drag, the effective diffusivity D^* of the ECS is reduced to only approximately 40% of that in free solution (D) [8].

1.3 The Solution: AAV as an Engineered Nanocarrier

Once the ECS is defined as a nanoscale porous medium, vector choice becomes a first-principles design problem: the carrier must:

(i) remain mobile within confined extracellular pathways

(ii) retain physical integrity in a crowded milieu

(iii) protect a genetic payload while enabling efficient neuronal transduction.



Figure 2: Schematic of the Adeno-Associated Virus (AAV) nanocarrier.

In this setting, Adeno-Associated Virus (AAV) provides a particularly well-matched architecture. As a non-enveloped particle, AAV lacks a lipid membrane and is instead built around a rigid protein capsid that protects the genome (the DNA that encodes the GCaMP family) during extracellular transit. Structurally, this capsid is icosahedral, a highly symmetric, 20-faceted shell assembled from 60 subunits, with an outer diameter of approximately 25 nm [5]. Even though technically AAV’s diameter is smaller than the 38–64 nm ECS “tunnels”, this clearance is in fact extremely tight; steric hindrance and complex electrostatic interactions with the extracellular matrix make actual in-tissue mobility non-trivial [6].

However, the porous-medium framework highlights a coupled engineering constraint: biological targeting is intrinsically linked to physical transport. Capsid features that promote cellular uptake, such as local charge patterns and receptor-binding motifs, can also increase nonspecific interactions with extracellular components, raising retention and reducing effective spread in tissue. Conversely, minimizing such interactions to maximize mobility may reduce engagement with cell-surface receptors and thereby decrease transduction efficiency. Consequently, capsid physicochemical properties, including surface charge distribution and hydrophobicity, must be co-optimized to balance tissue penetration against cellular entry in the nervous system [6].

Importantly, the AAV capsid also functions as a programmable molecular interface. Its surface topology and amino-acid sequence govern interactions with cell-surface receptors and glycans, effectively defining which cells it can enter, a property known as tropism [5]. This biological specificity is a powerful design lever in practice: it allows us to select or engineer capsid variants that are explicitly targeted to distinct neuronal populations. However, no sin-



Figure 3: Illustration of hindered diffusion of AAV nanocarriers within brain extracellular space (ECS). Cellular obstacles within this constrained porous medium force the ~ 25 nm viral particles to navigate highly tortuous pathways.

gle AAV serotype to date has fully resolved the fundamental trade-off between tissue penetration and cellular entry, making this an active engineering frontier [6].

2 PHOTOPHYSICAL TRANSDUCTION & SIGNAL SAMPLING

2.1 Calcium Imaging — The Sensor

Calcium imaging provides a "read" channel complementary to optogenetic perturbations: depolarization drives Ca^{2+} influx, and genetically encoded calcium indicators (GECIs, e.g. GCaMP) convert changes in free calcium concentration $[Ca^{2+}]$ into a fluorescence modulation reported as $\Delta F/F_0$ [11]. Figure 4 summarizes the minimal photophysical and biochemical ingredients that connect intracellular Ca^{2+} dynamics to detected photons.



Figure 4: Photophysical transduction in genetically encoded calcium indicators (schematic): (a) Jablonski diagram for fluorescence, (b) Ca^{2+}-dependent conformational switching that modulates non-radiative decay and quantum yield, and (c) illustrative Hill-type dose–response.

2.2 Fluorescence from first principles

Fluorescence in GFP-derived chromophores is an electronic transition within the singlet manifold. In a minimal Jablonski picture,

$$S_0 \xrightarrow{h\nu_{\mathrm{ex}}} S_1 \xrightarrow{\mathrm{IC}} S_1^{(v=0)} \xrightarrow{h\nu_{\mathrm{em}}} S_0 \tag{1}$$

where S_0 and S_1 are the electronic ground and first excited singlet states, ν_{ex} and ν_{em} are the excitation and emission frequencies, and IC denotes rapid internal conversion (vibrational relaxation) within S_1. Because energy is lost non-radiatively before emission, fluorescence is red-shifted relative to excitation. The Stokes shift is

$$\Delta\lambda \equiv \lambda_{\mathrm{em}} - \lambda_{\mathrm{ex}} > 0 \tag{2}$$

and physically reflects vibrational relaxation after absorption.

The fluorescence quantum yield Φ quantifies the probability that an absorbed excitation produces an emitted photon:

$$\Phi \equiv \frac{N_{\mathrm{em}}}{N_{\mathrm{abs}}} = \frac{k_r}{k_r + k_{nr}} \tag{3}$$

where k_r is the radiative decay rate and k_{nr} is the total non-radiative decay rate. Broadly, k_r is set by the transition dipole and optical environment, while k_{nr} is sensitive to chromophore flexibility and quenching pathways; in GFP-like proteins, rigidification typically suppresses k_{nr} and increases Φ.

2.3 Calcium-dependent conformational switching in GCaMP

GCaMP indicators combine circularly permuted GFP (cpGFP), calmodulin (CaM), and an M13 peptide [11]. When $[Ca^{2+}]$ rises, CaM binds Ca^{2+} ions cooperatively and engages M13, shifting the protein to a "closed" conformation that rigidifies the chromophore environment. The brightness increase is therefore primarily a change in Φ via suppressed non-radiative decay (reduced k_{nr} in Eq. (3)), not a large shift of the electronic gap $S_1 - S_0$.

Conformational dynamics and Ca^{2+} binding occur on μs–ms scales (and longer), whereas an excitation–emission cycle is $\sim$ ns; the conformation thus acts as a quasi-static parameter setting (k_r, k_{nr}).

The steady-state dose–response is commonly fit by a Hill form (Eq. (4)):

$$\frac{\Delta F/F_0}{(\Delta F/F_0)_{\max}} = \frac{[Ca^{2+}]^n}{K_d^n + [Ca^{2+}]^n} \tag{4}$$

where K_d (units: $\mathrm{mol\,L^{-1}}$) is an effective dissociation constant (half-activation) and n is the Hill coefficient. Typical

Variant	K_d (nM)	n	Source
GCaMP6s	144	2.9	[19]
jGCaMP7f	150	3.1	[14]

Table 1: Illustrative in vitro dose–response parameters for common GCaMP-family indicators. Reported values depend on conditions; they are included here only to set scales.

values are $K_d = \mathcal{O}(10^{-7}\,\mathrm{M})$ and $n \sim 3$; e.g. in vitro titrations report $K_d \approx 144\,\mathrm{nM}$ and $n \approx 2.9$ for GCaMP6s [19], while GENIE reports $K_d \approx 150\,\mathrm{nM}$ and $n \approx 3.1$ for jGCaMP7f [14].

2.4 SNR as a physical limit: photon statistics and two-photon brightness

The fluorescence trace is fundamentally a photon-counting measurement. If a time bin contains a baseline detected photon count F_0 (photons) and an activity-dependent increment ΔF, Poisson statistics give shot-noise standard deviation $\sqrt{F_0}$ and hence the shot-noise-limited signal-to-noise ratio (Eq. (5))

$$\mathrm{SNR} \equiv \frac{\Delta F}{\sqrt{F_0}}. \tag{5}$$

Under two-photon excitation (Section 3), the detected photon rate scales as (Eq. (6))

$$\dot{N}_{\mathrm{det}} \propto \eta\, \sigma_2\, \Phi\, C\, I^2 \tag{6}$$

where η is the net detection efficiency, σ_2 is the two-photon absorption cross-section (often quoted in Göeppert–Mayer units, $1\,\mathrm{GM} = 10^{-58}\,\mathrm{m}^4\,\mathrm{s\,photon}^{-1}$), C is indicator concentration, and I is the focal intensity [16, 21]. Thus σ_2 and Φ are molecular levers on the shot-noise-limited SNR, alongside optical collection.

A practical sensitivity advantage of cytosolic calcium indicators over membrane-localized voltage reporters is geometric: for a soma radius $r \sim 10\,\mu\mathrm{m}$ and membrane thickness $\delta \sim 5\,\mathrm{nm}$, $V/(A\delta) \sim r/(3\delta) \approx 10^2$–$10^3$, supporting many more emitters and hence larger F_0 (with caveats: membrane density and (σ_2, Φ) differ by sensor).

2.5 Kinetics and temporal resolution (and the correct role of sampling theory)

Calcium imaging does not aim to reconstruct membrane voltage. An action potential has FWHM $\sim 1\,\mathrm{ms}$ (bandwidth $\sim 1\,\mathrm{kHz}$), whereas typical raster-scanned two-photon imaging runs at $\sim 30\,\mathrm{Hz}$ frame rate [18]. The measured observable is instead the low-pass-filtered calcium/indicator dynamics with correlation time τ; its relevant bandwidth is $B_{\mathrm{eff}} \sim (2\pi\tau)^{-1}$, so Nyquist sampling can be satisfied for calcium transients even though it is not for voltage.

A minimal kinetic model treats binding as a reversible reaction,

$$\mathrm{GCaMP} + n\,\mathrm{Ca}^{2+} \underset{k_{\mathrm{off}}}{\overset{k_{\mathrm{on}}}{\rightleftharpoons}} \mathrm{GCaMP}\cdot\mathrm{Ca}_n \tag{7}$$

with effective association rate k_{on} and dissociation rate k_{off}. Linearizing an occupancy variable around a working point gives a relaxation rate $\tau^{-1} \approx k_{\mathrm{on}}[\mathrm{Ca}^{2+}]^n + k_{\mathrm{off}}$, implying rise/decay times in the ~ 10–$10^3\,\mathrm{ms}$ range in neurons [11]. This temporal blurring is a biophysical constraint that limits spike timing precision and motivates continued engineering (e.g., GCaMP6f and the jGCaMP8 family), not a deliberate anti-aliasing feature [11, 20].

3 SCATTERING MITIGATION AND TWO-PHOTON READOUT

3.1 The Readout: Overcoming the Scattering Channel

Having established the physical mechanisms for delivering the genetic cargo and transducing electrical activity into optical signals, we now confront the fundamental challenge of extracting this information from brain tissue. The brain is a highly scattering medium where photons undergo random walks due to inhomogeneous refractive index distributions across cellular membranes, organelles, and extracellular structures. This scattering destroys spatial localization, the very information we need in order to resolve individual neurons. In this section, we demonstrate why conventional one-photon microscopy fails in this regime and how two-photon excitation provides a physics-based solution through non-linear optical confinement [22].



Figure 5: Two-photon readout in scattering tissue (schematic): (a) axial confinement from nonlinear excitation (I^2), (b) reduced scattering at longer wavelengths (illustrative power-law trend), and (c) excitation geometry comparison for one-photon and two-photon microscopy.

3.2 The Noise Channel: Scattering as a Spatial Filter

When light propagates through brain tissue, it encounters refractive index variations on the order of $\Delta n \approx 0.05$ at cellular boundaries. These discontinuities cause both absorption and elastic scattering. The attenuation follows the Beer-Lambert

Law:

$$I(z) = I_0 \exp(-\mu_t z) \tag{8}$$

where $\mu_t = \mu_a + \mu_s$ is the total attenuation coefficient comprising absorption (μ_a) and scattering (μ_s) components. In biological tissue, scattering dominates: $\mu_s \gg \mu_a$ [16]. The characteristic penetration depth, or mean free path (MFP), is defined as $l_s = 1/\mu_s$. For visible light ($\sim 500\,\text{nm}$) in cortical tissue, $l_s \approx 50\,\mu\text{m}$, far shorter than the depth of neuronal populations of interest (hundreds of microns).

The wavelength-dependence of scattering is governed by Rayleigh scattering for particles much smaller than the wavelength ($d \ll \lambda$), where the scattering cross-section scales as:

$$\sigma_s \propto \lambda^{-4} \tag{9}$$

This strong wavelength dependence has critical implications: longer wavelengths scatter less [25] [26]. By shifting from visible light (500 nm) to near-infrared (NIR) wavelengths around 900 nm, we can increase the mean free path by approximately $(900/500)^4 \approx 10\times$, extending l_s to $\sim 500\,\mu\text{m}$.

Shorter wavelengths scatter more (reducing penetration), while longer wavelengths fall outside the absorption spectrum of typical fluorescent proteins (green fluorescent protein absorbs at $\sim 488\,\text{nm}$)[16]. Two-photon excitation elegantly resolves this by using NIR photons for excitation while still targeting the visible absorption band through simultaneous two-photon absorption.

The consequence of scattering is the loss of spatial coherence. Photons emitted from a fluorophore deep in tissue undergo multiple scattering events before reaching the detector, randomizing their direction. In conventional (confocal) one-photon microscopy, excitation light forms a cone illuminating not just the focal point but the entire depth of the sample. All fluorophores within this cone emit, creating a high background signal. While confocal systems use a spatial pinhole to reject out-of-focus light, scattered photons from the excitation cone still reach the pinhole, degrading the signal-to-noise ratio (SNR) and limiting depth penetration to $\sim 100\,\mu\text{m}$ in scattering tissue. [23]

3.3 The Spatial Filter: Non-linear Optical Confinement

Two-photon microscopy solves the scattering problem through a fundamentally different physical mechanism: non-linear excitation. The critical distinction lies in the probability of excitation as a function of light intensity.

Linear (One-Photon) Excitation. In conventional fluorescence, a single photon with energy $E = h\nu$ matching the energy gap between ground state S_0 and excited state S_1 is absorbed. The excitation rate depends linearly on photon flux (intensity I):

$$P_1 \propto I \tag{10}$$

Because excitation is linear, it occurs anywhere the light is present; throughout the entire illumination cone. Even with tight focusing, fluorophores above and below the focal plane are excited proportionally to the local intensity.

Non-linear (Two-Photon) Excitation. Two-photon absorption requires the quasi-simultaneous arrival of two photons, each with approximately half the energy ($E/2 = h\nu/2$), within a temporal window set by the uncertainty principle ($\Delta t \approx 10^{-15}$ s). The excitation proceeds via a virtual state, an intermediate, non-stationary quantum state that exists only during the absorption process. Because the probability of two photons arriving simultaneously at a given molecule is proportional to the square of the photon density, the excitation rate scales quadratically [22]:

$$P_2 \propto I^2 \tag{11}$$

This quadratic dependence has profound spatial consequences. For a Gaussian beam focused to a diffraction-limited spot, the intensity profile is:

$$I(r) = I_0 \exp\left(-\frac{2r^2}{w_0^2}\right) \tag{12}$$

where w_0 is the beam cross-sectional radius. Squaring this intensity ($P_2 \propto I^2$) yields:

$$P_2(r) \propto \exp\left(-\frac{4r^2}{w_0^2}\right) \tag{13}$$

The excitation volume shrinks by a factor of $\sqrt{2}$ compared to one-photon excitation, meaning the probability distribution is narrower. More importantly, the excitation decays much faster with distance from the focal plane. At a distance $r = w_0$ (one beam waist), one-photon excitation has decayed to $e^{-2} \approx 13.5\%$ of peak, while two-photon excitation has decayed to $e^{-4} \approx 1.8\%$; an order of magnitude difference in suppression of out-of-focus excitation [21].

This non-linear confinement provides inherent optical sectioning. Only fluorophores within the focal volume (where photon density is highest) experience significant excitation. Fluorophores above or below the focal plane see dramatically reduced intensity and thus exponentially reduced excitation probability. Critically, this localization occurs *before* photon emission, making it immune to scattering. Even if emitted photons scatter on their way to the detector, they originated from a well-defined spatial location (the focal volume), preserving spatial information. This is the fundamental advantage: two-photon microscopy acts as a virtual spatial filter.

To achieve the required photon densities for two-photon absorption (I^2 scaling), we use mode-locked femtosecond lasers (pulse width $\sim 100\,\mathrm{fs}$, repetition rate $\sim 80\,\mathrm{MHz}$) [21]. These ultrashort pulses concentrate photon flux in time, increasing the instantaneous intensity by $\sim 10^5$ compared to continuous-wave illumination at the same average power. The peak intensity at the focal point can exceed $10^{29}\,\mathrm{photons}/(\mathrm{cm}^2 \cdot \mathrm{s})$, sufficient for detectable two-photon absorption while maintaining average powers low enough ($\sim 10\,\mathrm{mW}$) to avoid photodamage.

The practical consequence is unparalleled depth penetration. By combining long-wavelength excitation (reduced scattering) with non-linear spatial confinement (elimination of out-of-focus background), two-photon microscopy routinely images through 500–800 μm [16] of cortical tissue, depths inaccessible to confocal microscopy. The detected signal, while reduced in absolute magnitude due to scattering losses, maintains high contrast because all photons detected originated from the focal volume. The brain's scattering medium, previously an insurmountable barrier, becomes merely an inefficient but navigable transmission channel.

In summary, two-photon excitation solves the scattering problem by exploiting the I^2 dependence of non-linear absorption. This quadratic scaling concentrates excitation to a diffraction-limited focal volume, providing intrinsic optical sectioning [22] without physical pinholes. The combination of NIR wavelengths (minimizing scattering via λ^{-4} scaling) and non-linear confinement (rejecting out-of-focus fluorescence) enables deep-tissue imaging of neural activity.

4 OUTLOOK: NEURAL DYNAMICS

Viral-mediated two-photon imaging yields $\Delta F/F_0$ traces for many neurons (Sections 2 and 3). For N recorded neurons, the measured population state at time t is a point $\mathbf{x}(t) \in \mathbb{R}^N$ with components $x_i(t) = \Delta F_i(t)/F_{0,i}$.

4.1 From fluorescence traces to a low-dimensional state space

Let $\mathbf{X} \in \mathbb{R}^{T\times N}$ be the data matrix of mean-subtracted traces sampled at T time points. Principal Component Analysis (PCA) diagonalizes the covariance

$$\mathbf{C} = \frac{1}{T}\mathbf{X}^\top\mathbf{X} \tag{14}$$

yielding eigenvalues $\{\lambda_k\}$ and eigenvectors (principal components) that define orthogonal collective coordinates [12]. Empirically, the spectrum often decays rapidly, motivating an effective dimension $d_{\mathrm{eff}} \ll N$ defined operationally (e.g., explained-variance threshold or participation ratio),

$$d_{\mathrm{eff}} \equiv \frac{(\sum_k \lambda_k)^2}{\sum_k \lambda_k^2}. \tag{15}$$

Correlated firing reduces independent degrees of freedom, so trajectories concentrate near a low-dimensional "neural manifold" embedded in $\mathbb{R}^N$ [10, 15].

Crucially, PCA is applied to fluorescence traces rather than spikes. The measurement operator discussed in Section 2 (indicator kinetics) and Section 3 (optical readout and photon statistics) shapes both the covariance spectrum and any inferred geometry. Figure 6 illustrates this point with a toy model: even when the latent dynamics are fixed, stronger temporal filtering and lower SNR can change the apparent eigen-spectrum and hence the estimated d_{eff}.



Figure 6: Illustrative toy-model validation for PCA-based effective dimensionality. A low-dimensional latent trajectory is mixed into N neurons, converted to spikes, filtered by an exponential calcium kernel with time constant τ, and corrupted by measurement noise. Increasing τ and/or reducing SNR changes the covariance eigen-spectrum and the participation-ratio estimate of d_{eff}, even when the latent dynamics are unchanged.

4.2 Attractors, stability, and the measurement operator

Within this reduced state space, fixed points, limit cycles, and more complex attractors correspond to reproducible population activity patterns, with stimuli or behavioural contexts selecting different basins of attraction [17]. Local stability can, in principle, be quantified via a Lyapunov exponent estimated from trajectories.

The calcium sensor itself shapes these trajectories. In minimal form, the measured fluorescence is a convolution of an underlying spike-driven signal $s(t)$ with an impulse response $h(t)$: $x(t) \approx (s * h)(t) + \eta(t)$, with η dominated by photon noise. Since $H(\omega)$ suppresses high-frequency components, this filtering can reduce apparent variability but also smooth fast transitions and bias inferred dimensionality; quantitative interpretations benefit from forward models that include the sensor as part of the measurement operator.

APPLICATION & CONCLUSION

We have presented a complete physical pipeline in which each stage imposes quantifiable constraints on the next: AAV nanocarriers navigate a porous medium whose tortuosity and electrostatic interactions limit transduction efficiency; GECIs low-pass filter the resulting calcium transients with kinetics that bound temporal resolution; two-photon excitation provides the spatial confinement needed to resolve single neurons despite tissue scattering; and dimensionality reduction extracts latent dynamics from traces shaped by all preceding measurement operators. Crucially, these stages are not independent, the SNR delivered by the optical system sets the floor on the reliability of PCA-inferred manifold geometry, and the kinetic bandwidth of the GECI determines whether fast transitions in attractor state can be resolved at all. Understanding the pipeline as a coupled system, rather than a sequence of isolated techniques, is therefore essential for interpreting any inferred neural dynamics quantitatively.

This framing exposes the most pressing current limitations. On the delivery side, no AAV serotype has fully resolved the trade-off between tissue penetration and cellular specificity identified in Section 1, meaning population coverage remains incomplete and potentially biased toward accessible cell types [6]. On the transduction side, even the fastest current indicators — the jGCaMP8 family, with decay times approaching 50 ms [20], cannot faithfully resolve individual spikes at physiological firing rates above 20 Hz without deconvolution algorithms that themselves introduce assumptions about spike statistics. This means that inferred attractor trajectories in Section 4 are filtered representations of the underlying dynamics, and quantitative claims about dimensionality or Lyapunov stability require forward models that explicitly account for the sensor's impulse response.

Looking forward, each bottleneck identified here is an active engineering frontier. Chemogenetic and intersectional viral strategies promise more precise cell-type targeting [5, 6]; protein engineering of GECIs continues to push Kd and kinetic parameters toward the single-spike detection regime [20]; and emerging volumetric two-photon and light-field approaches aim to extend the field of view from hundreds to thousands of simultaneously recorded neurons without sacrificing the spatial confinement that makes deep-tissue imaging possible [16]. On the inference side, methods that jointly model the measurement operator and the latent dynamics, rather than treating fluorescence traces as direct proxies for activity, are beginning to correct for the biases identified in Section 4 [12].

A recent study illustrates how this pipeline can be extended beyond neural activity itself: Fernández-Ballester et al. [9] used viral vectors to deliver fluorescent probes targeting the extracellular matrix of a mouse brain, then applied two-photon imaging to capture real-time structural dynamics of the brain microenvironment simultaneously with calcium signals. This demonstrates that the physical framework developed here generalises naturally to multi-channel imaging of the full neural niche, not merely the neurons within it.

Viral-mediated two-photon calcium imaging is therefore best understood not as a finished technique but as an emerging field and physically principled measurement system whose ultimate resolution; in space, time, and neural population, is set by tractable biophysical constraints. Progress on each of these constraints, guided by the kind of first-principles analysis presented here, is what will determine how completely we can decode the brain's low-dimensional dynamics from the photons we are able to collect and eventually further understand the human brain and its processes.

References

[1] Grienberger, C. and Konnerth, A. (2012). Imaging calcium in neurons. *Neuron*, 73(5), 862–885.

[2] Peron, S., Chen, T.-W. and Svoboda, K. (2015). Comprehensive imaging of Cortical Networks. *Current Opinion in Neurobiology*, 32, 115–123.

[3] Syková, E. and Nicholson, C. (2008). Diffusion in brain extracellular space. *Physiological Reviews*, 88(4), 1277–1340.

[4] Thorne, R.G. and Nicholson, C. (2006). In vivo diffusion analysis with quantum dots and Dextrans predicts the width of brain extracellular space. *Proceedings of the National Academy of Sciences*, 103(14), 5567–5572.

[5] Naso, M.F. et al. (2017). Adeno-associated virus (AAV) as a vector for gene therapy. *BioDrugs*, 31(4), 317–334.

[6] Hudry, E. and Vandenberghe, L.H. (2019). Therapeutic AAV gene transfer to the nervous system: A clinical reality. *Neuron*, 102(1), 263–277.

[7] Tønnesen, J., Hrabětová, S. and Soria, F.N. (2023). Local diffusion in the extracellular space of the brain. *Neurobiology of Disease*, 177, 105981.

[8] Thorne, R.G. et al. (2008). In vivo diffusion of lactoferrin in brain extracellular space is regulated by interactions with heparan sulfate. *Proceedings of the National Academy of Sciences*, 105(24), 8416–8421.

[9] Fernández-Ballester, M. et al. (2025). Viral-mediated fluorescent labeling of hyaluronan reveals extracellular matrix dynamics in the Mouse Brain in vivo. *bioRxiv* (Preprint).

[10] Chaudhuri, R., He, B.-J., & Wang, X.-J. (2019). The Geometry and Dimensionality of Brain-wide Activity. *Nature*, **572**, 61–65.

[11] Chen, T.-W., Wardill, T. J., Sun, Y., et al. (2013). Ultrasen-

sitive fluorescent proteins for imaging neuronal activity. *Nature*, **499**, 295–300.

[12] Cunningham, J. P., & Yu, B. M. (2014). Dimensionality reduction for large-scale neural recordings. *Nature Neuroscience*, **17**(11), 1500–1509.

[13] Dana, H., Sun, Y., Mohar, B., et al. (2019). High-performance calcium sensors for imaging activity in neuronal populations and microcompartments. *Nature Methods*, **16**, 649–657.

[14] GENIE Project Team. (2021). *jGCaMP8 calcium indicators* (online resource; includes jGCaMP7f purified-protein parameters). HHMI Janelia Research Campus.

[15] Gallego, J. A., Perich, M. G., Miller, L. E., & Solla, S. A. (2017). Neural manifolds for the control of movement. *Neuron*, **94**(5), 978–984.

[16] Helmchen, F., & Denk, W. (2005). Deep tissue two-photon microscopy. *Nature Methods*, **2**(12), 932–940.

[17] Khona, M., & Fiete, I. R. (2022). Attractor and integrator networks in the brain. *Nature Reviews Neuroscience*, **23**, 744–766.

[18] Pologruto, T. A., Sabatini, B. L., & Svoboda, K. (2003). ScanImage: Flexible software for operating laser scanning microscopes. *BioMedical Engineering OnLine*, **2**, 13.

[19] Wu, N., Nishioka, W. K., Derecki, N. C., & Maher, M. P. (2019). High-throughput-compatible assays using a genetically-encoded calcium indicator. *Scientific Reports*, **9**, 12692.

[20] Zhang, Y., Rózsa, M., Liang, Y., et al. (2023). Fast and sensitive GCaMP calcium indicators for imaging neural populations. *Nature*, **615**, 884–891.

[21] Zipfel, W. R., Williams, R. M., & Webb, W. W. (2003). Nonlinear magic: multiphoton microscopy in the biosciences. *Nature Biotechnology*, **21**(11), 1369–1377.

[22] Denk, W., Strickler, J.H. and Webb, W.W. (1990). Two-photon laser scanning fluorescence microscopy. *Science*, 248(4951), 73–76.

[23] Centonze, V.E. and White, J.G. (1998). Multiphoton excitation provides optical sections from deeper within scattering specimens than confocal imaging. *Biophysical Journal*, 75(4), 2015–2024.

[24] Theer, P., Hasan, M.T. and Denk, W. (2003). Two-photon imaging to a depth of 1000 μm in living brains by use of a Ti:Al_2O_3 regenerative amplifier. *Optics Letters*, 28(12), 1022–1024.

[25] Jacques, S.L. (2013). Optical properties of biological tissues: a review. *Physics in Medicine & Biology*, 58(11), R37–R61.

[26] Tuchin, V.V. (2015). *Tissue Optics: Light Scattering Methods and Instruments for Medical Diagnosis* (3rd ed.). SPIE Press.