

CE CONFORMITY DECLARATION
93/42/CEE, as modified by the amendment 2007/47/CE

Referenced to the Medical Device

"EASYLAX NEW.FA.DEM. MICROCLISMA GR. 3 BAMBINI" code 70014 and with the following commercial variant:

- "Pomoli 3g" code 70479
- "Eurospital 3g" code 70469
- "Glicerol Inteli 3g" code 70502
- "Microcell 3g" code 70608
- "Assista 3g" code 70609
- "Sintofarm 3g" code 70611
- "4Lax 3g" code 70613
- "Milax-m 3g" code 70615
- "Osmoforce Lax G 3g" code 70626
- "FertoLax 3g" code 70631
- "Laxik 3g" code 70639

Giugliano in Campania (NA), 12.02.2024

The company New.Fa.Dem. Srl, represented by Maria Goncharova (CEO) with legal office in Giugliano in Campania (NA), Viale Ferrovie dello Stato, 1 – Zona ASI 80014, declare under his own responsibility to be the manufacturer of the medical devices "EASYLAX NEW.FA.DEM. MICROCLISMA GR. 3 BAMBINI", "Pomoli 3g", "Eurospital 3g", "Glicerol Inteli 3g", "Microcell 3g", "Assista 3g", "Sintofarm 3g", "4Lax 3g", "Milax-m 3g", "Osmoforce Lax G 3g", "FertoLax 3g" and "Laxik 3g"

It is also declared that:

- We herewith declare that the above mentioned product meet the provisions of the Council Directive 93/42/EEC and 2007/47/EEC for Medical Device;
- The present medical device belongs to the class I according to the rules 5 of the Annex IX – Amendment 93/42/EEC (introduced in Italy with the Legislative Bill dated 24 February 1997, nr. 46 and succeeding modifications);
- The present medical device did not include any substance, which, if used separately, could be considered a medicine according to the art. 1 of the Bill 65/65/EEC;
- The present medical device is commercialized as NOT STERILE;
- The related medical device IS NOT A MEASURE INSTRUMENT;
- The related medical device IS NOT DESTINATED FOR CLINICAL INVESTIGATION;
- The manufacturer observed all dispositions applicable of the annex I and VII of Directive 93/42/EEC.

Harmonized Rules in compliance with the declaration:

UNI CEI EN ISO 14971:2012 *Application of risk management to medical devices.*

UNI CEI EN 1041:2013 *Information supplied by the manufacturer of medical devices. (Non-harmonized standard – the harmonized standard is previous version of 2008).*


UNI CEI EN ISO 15223-1:2017 *Medical devices - Symbols to be used with medical device. labels, labelling and information to be supplied - Part 1: General requirements.*

UNI CEI EN ISO 13485:2016 - *Medical devices -Quality management systems.*

UNI EN ISO 10993-1:2010 - *Biological evaluation of medical devices - Part 1:Evaluation and testing within a risk management process.*

This Conformity Declaration is valid only for above mentioned devices and used in a proper way as specified in the technical and security data sheet.

New.Fa.Dem. s.r.l.
Zona Ind. ASI - Via Ferrovia dello Stato, 1
80014 Giugliano in Campania (NA) Italy
T. +39 (0) 081.8195250 • +39 (0) 081.9757469
F. +39 081.8186893
P.I. 01499511218 • C.F. 06115290634
www.newfadem.it



New. Fa.Dem. S.r.l.
Maria Goncharova