



Work Order Report

Asset Information

Manufacturer: InterMed Ltd

Model: HMP 2070-40HC

Serial Number: 10_2014-517

Work Order Details

Work Order #: 1

Description: Responded to service request regarding reported loud noise from pasteurizer pump during pasteurization cycle. Coordinated with staff to ensure the unit was available for testing. Simulated operation by manually filling unit and loading bottles to advance to pasteurization cycle. Observed both fill and circulation pumps during heating and pasteurization. No abnormal noises were detected throughout testing. Advised staff to monitor and document time/conditions if noise recurs.

Status: In Progress

Assigned To: admin

Scheduled Date: 8/12/2025

Completion Date: N/A

