

HACH COMPANY



An ISO 9001 Certified Company

P.O.Box 389
Loveland, CO 80539
(970) 669-3050

Certificate of Analysis

Page 1

COMMODITY: **STABLCAL® FORMAZIN STANDARD 1000 NTU**
COMMODITY NUMBER: **2660601** MANUFACTURE DATE: **7/14/2020**
LOT NUMBER: **A0177**

DATE OF ANALYSIS: **7/23/2020**

<i>TEST</i>	<i>SPECIFICATIONS</i>	<i>RESULTS</i>
Turbidity	950 to 1050 NTU	1031.0 NTU

The expiration date is Oct 2021

Formazin and StablCal® solutions provided by Hach are not NIST traceable because the NIST does not carry turbidity standards. However, the use of Formazin and StablCal® as used in Hach method 8195 are accepted by the EPA as a primary standard to be used in the calibration of turbidity instruments.

APROVADO
Responsável: <u>Wellington</u>
Padrão: <u>2660601</u>
Data: <u>14/07/2020</u>
Validade: <u>10/2021</u>

Certified by _____

Scott Als

Scott Als
Analytical Services Chemist