

P.O.Box 389 Loveland, CO 80539 (970) 669-3050

Certificate of Analysis
This is a Component of 2659505-BR lot A1355

Page 1

COMMODITY: STABLCAL | sup | TS | sup 0 FORMAZIN STANDARD

1000 NTU

DATE OF ANALYSIS:

COMMODITY NUMBER: 2660601

MANUFACTURE DATE:

12/17/2021

LOT NUMBER: A1326

12/14/2021

TEST

SPECIFICATIONS

RESULTS

Turbidity

950 to 1050 NTU

1016.0 NTU

The expiration date is Mar 2023

Formazin and StablCal® solutions provided by Hach are not NIST traceable because the NIS does not carry turbidity standards. However, the use of Formazin and StablCal® as use in Hach method 8195 are accepted by the EPA as a primary standard to be used in the calibration of turbidity instruments.

APROVADO

Responsável: RENAMO

Padrão: Z6606.01

Data: 02/06/2022

Validade: 30/03/2023

Certified by _

Scott Als

Analytical Services Chemist

east als