

Certificate of Analysis List

For request number 828010

Catalog Number Entered	Lot Number Entered	Related Catalog Number	Related Lot Code	Description	
2660501	6190	N/A	N/A	STABLCAL [®] FORMAZIN STANDARD	800
NTU					

Total Enclosures: 1

*Certificate of Analysis*

Page 1

COMMODITY: **STABLCAL® FORMAZIN STANDARD 800 NTU**
COMMODITY NUMBER: **26605-01** MANUFACTURE DATE:
LOT NUMBER: **A6190** **7/14/2016**

DATE OF ANALYSIS:
7/19/2016

<i>TEST</i>	<i>SPECIFICATIONS</i>	<i>RESULTS</i>
Turbidity	760 to 840 NTU	813.9 NTU

The expiration date is Oct 2017

Formazin and StablCal® solutions provided by Hach are not NIST traceable because the NIST does not carry turbidity standards. However, the use of Formazin and StablCal® as used in Hach method 8195 are accepted by the EPA as a primary standard to be used in the calibration of turbidity instruments.

Certified by _____

Scott Als
Analytical Services Chemist