

**Certificate of Analysis**

Page 1

COMMODITY: **StablCal|sup|TS|sup0 Standard, 1.0 NTU**COMMODITY NUMBER: **2659842**

MANUFACTURE DATE:

LOT NUMBER: **A10334****12/4/2022**

DATE OF ANALYSIS:

12/4/2022

TEST	SPECIFICATIONS	RESULTS
Turbidity	0.95 to 1.05 NTU	1.030 NTU

The expiration date is Nov 2023

Formazin and StablCal® solutions provided by Hach are not NIST traceable because the NIST does not carry turbidity standards. However, the use of Formazin and StablCal® as used in Hach method 8195 are accepted by the EPA as a primary standard to be used in the calibration of turbidity instruments.

APROVADO
Responsável: <u>RENATO</u>
Padrão: <u>26598.42</u>
Data: <u>02/06/2022</u>
Validade: <u>30/11/2023</u>

Certified by

Scott Als
Analytical Services Chemist