

**Certificate of Analysis****This is a Component of 2971205 lot A10032**

Page 1

COMMODITY: **STABLCAL|sup|TS|sup0 FORMAZIN STANDARD 100 NTU**COMMODITY NUMBER: **2684901**

MANUFACTURE DATE:

DATE OF ANALYSIS:

LOT NUMBER: A10010

1/15/2022**1/19/2022**

TEST	SPECIFICATIONS	RESULTS
Turbidity	95 to 105 NTU	102.3 NTU

The expiration date is Apr 2023

Formazin and StablCal® solutions provided by Hach are not NIST traceable because the NIST does not carry turbidity standards. However, the use of Formazin and StablCal® as used in Hach method 8195 are accepted by the EPA as a primary standard to be used in the calibration of turbidity instruments.

APROVADO
Responsável: <u>RENATO</u>
Padrão: <u>26849.01</u>
Data: <u>02/06/2022</u>
Validade: <u>30/04/2023</u>

Certified by _____

Scott Als
Analytical Services Chemist