

HACH COMPANY



An ISO 9001 Certified Company

P.O.Box 389  
Loveland, CO 80539  
(970) 669-3050

**Certificate of Analysis**

**This is a Component of 2971205-BR lot A2180**

Page 1

COMMODITY: **STABLCAL|sup|TS|sup0 FORMAZIN STANDARD 100 NTU**

COMMODITY NUMBER: **2684901**

MANUFACTURE DATE:

DATE OF ANALYSIS:

LOT NUMBER: **A2139**

**6/7/2022**

**6/14/2022**

TEST	SPECIFICATIONS	RESULTS
Turbidity	95 to 105 NTU	101.0 NTU

The expiration date is Sep 2023

Formazin and StablCal® solutions provided by Hach are not NIST traceable because the NIST does not carry turbidity standards. However, the use of Formazin and StablCal® as used in Hach method 8195 are accepted by the EPA as a primary standard to be used in the calibration of turbidity instruments.

<b>APROVADO</b>
Responsável: <u>RENATO</u>
Padrão: <u>26849.01</u>
Data: <u>13/10/2022</u>
Validade: <u>30/09/2023</u>

Certified by

A handwritten signature in cursive script that reads "Scott Als".

Scott Als  
Analytical Services Chemist