

***Certificate of Analysis******This is a Component of 2971205 lot A0183***

Page 1

COMMODITY: **STABLCAL® FORMAZIN STANDARD 800 NTU**COMMODITY NUMBER: **2660501**

MANUFACTURE DATE:

DATE OF ANALYSIS:

LOT NUMBER: **A0169****6/23/2020****6/28/2020**

---

<b>TEST</b>	<b>SPECIFICATIONS</b>	<b>RESULTS</b>
Turbidity	760 to 840 NTU	817.0 NTU

The expiration date is Sep 2021

Formazin and StablCal® solutions provided by Hach are not NIST traceable because the NIS does not carry turbidity standards. However, the use of Formazin and StablCal® as use in Hach method 8195 are accepted by the EPA as a primary standard to be used in the calibration of turbidity instruments.

<b>APROVADO</b>
Responsável: <u>RENATO</u>
Padrão: <u>26605.01 - C14</u>
Data: <u>21/09/2020</u>
Validade: <u>30/09/2021</u>

Certified by

Scott Als  
Analytical Services Chemist