

# Certificate of Analysis List

For request number 828009

Catalog Number Entered	Lot Number Entered	Related Catalog Number	Related Lot Code	Description	
2684901 NTU	6190	N/A	N/A	STABLCAL <sup>®</sup> FORMAZIN STANDARD	100

Total Enclosures: 1

*Certificate of Analysis*

Page 1

COMMODITY: **STABLCAL® FORMAZIN STANDARD 100 NTU**  
COMMODITY NUMBER: **26849-01** MANUFACTURE DATE:  
LOT NUMBER: **A6190** **7/15/2016**

DATE OF ANALYSIS:  
**7/19/2016**

---

<i><b>TEST</b></i>	<i><b>SPECIFICATIONS</b></i>	<i><b>RESULTS</b></i>
Turbidity	95 to 105 NTU	102.5 NTU

The expiration date is Oct 2017

Formazin and StablCal® solutions provided by Hach are not NIST traceable because the NIST does not carry turbidity standards. However, the use of Formazin and StablCal® as used in Hach method 8195 are accepted by the EPA as a primary standard to be used in the calibration of turbidity instruments.

Certified by \_\_\_\_\_

Scott Als  
Analytical Services Chemist