

*Certificate of Analysis*

Page 1

COMMODITY: **STABLCAL^{sup} | TS^{sup} | 0 FORMAZIN STANDARD 200 NTU**COMMODITY NUMBER: **2660401**

MANUFACTURE DATE:

DATE OF ANALYSIS:

LOT NUMBER: **A2273****10/24/2022****10/31/2022**

<i>TEST</i>	<i>SPECIFICATIONS</i>	<i>RESULTS</i>
Turbidity	190 to 210 NTU	202.0 NTU

The expiration date is Jan 2024

Formazin and StablCal® solutions provided by Hach are not NIST traceable because the NIS does not carry turbidity standards. However, the use of Formazin and StablCal® as use in Hach method 8195 are accepted by the EPA as a primary standard to be used in the calibration of turbidity instruments.

APROVADO
Responsável: <u>RENATO</u>
Padrão: <u>26604.01</u>
Data: <u>03/12/2022</u>
Validade: <u>31/01/2024</u>

Certified by

Scott Als
Analytical Services Chemist