



An ISO 9001 Certified Company

Certificate of Analysis**This is a Component of 2659505-BR lot A1355**

Page 1

COMMODITY: **STABLCAL|sup|TS|sup0 FORMAZIN STANDARD 200 NTU**COMMODITY NUMBER: **2660401**

MANUFACTURE DATE:

DATE OF ANALYSIS:

LOT NUMBER: **A1300****11/23/2021****11/30/2021**

TEST	SPECIFICATIONS	RESULTS
Turbidity	190 to 210 NTU	206.0 NTU

The expiration date is Feb 2023

Formazin and StablCal® solutions provided by Hach are not NIST traceable because the NIS does not carry turbidity standards. However, the use of Formazin and StablCal® as use in Hach method 8195 are accepted by the EPA as a primary standard to be used in the calibration of turbidity instruments.

APROVADO
Responsável: <u>RENATO</u>
Padrão: <u>26604.01</u>
Data: <u>02/06/2022</u>
Validade: <u>28/02/2023</u>

Certified by

Scott Als
Analytical Services Chemist