

*Certificate of Analysis*

Page 1

COMMODITY: **STABLCAL|sup|TS|sup0 FORMAZIN STANDARD** 7500 NTU
COMMODITY NUMBER: **2584201** MANUFACTURE DATE: DATE OF ANALYSIS:
LOT NUMBER: **A2264** 10/3/2022 10/27/2022

TEST	SPECIFICATIONS	RESULTS
Formazine Concentration	7125 to 7875 NTU	7306.3 NTU

The expiration date is Jan 2024

APROVADO	
Responsável:	RENATO
Padrão:	25842.01
Data:	03/12/2022
Validade:	31/01/2024

Certified by

Scott Als
Analytical Services Chemist