

Certificate of Analysis List

For request number 794643

Catalog Number Entered	Lot Number Entered	Related Catalog Number	Related Lot Code	Description
2370853	5097	N/A	N/A	Chloride Standard Solution 0.00282 N

Total Enclosures: 1

*Certificate of Analysis*

Page 1

COMMODITY: **Chloride Standard Solution 0.00282 N**COMMODITY NUMBER: **23708-53**

MANUFACTURE DATE:

LOT NUMBER: **A5097****4/6/2015**

DATE OF ANALYSIS:

4/6/2015

<i>TEST</i>	<i>SPECIFICATIONS</i>	<i>RESULTS</i>
Sodium Chloride Concentration in mg/L	99 to 101 mg/L	100.9 mg/L
Sodium Chloride Concentration in Normality	0.00279 to 0.00285	0.002845

The expiration date is Apr 2020

The item 2370853 is traceable to NIST standards SRM 2201 Sodium Chloride LOT N/A.

Certified by

Scott Als
Analytical Services Chemist