

HACH COMPANY



An ISO 9001 Certified Company

P.O.Box 389  
Loveland, CO 80539  
(970) 669-3050

### Certificate of Analysis

Page 1

COMMODITY: **StablCal|sup|TS|sup** Solution, <0.1 NTU  
COMMODITY NUMBER: **2659742-BR** MANUFACTURE DATE:  
LOT NUMBER: **A1194** 7/23/2021

DATE OF ANALYSIS:  
7/23/2021

TEST	SPECIFICATIONS	RESULTS
Turbidity of diluent	0 to 0.1 NTU	0.07 NTU

The expiration date is Jul 2023

<b>APROVADO</b>
Responsável: <u>Wellington</u>
Padrão: <u>MRC1 - 0,1 NTU</u>
Data: <u>23/07/2021</u>
Validade: <u>30/07/2023</u>

Certified by

A handwritten signature in cursive script that reads "Scott Als".

Scott Als  
Analytical Services Chemist