

Certificate of Analysis List

For request number 828014

Catalog Number Entered	Lot Number Entered	Related Catalog Number	Related Lot Code	Description	
2684801	6193	N/A	N/A	STABLCAL [®] FORMAZIN STANDARD	20
NTU					

Total Enclosures: 1

*Certificate of Analysis*

Page 1

COMMODITY: **STABLCAL® FORMAZIN STANDARD 20 NTU**
COMMODITY NUMBER: **26848-01** MANUFACTURE DATE:
LOT NUMBER: **A6193** **7/18/2016**

DATE OF ANALYSIS:
7/22/2016

<i>TEST</i>	<i>SPECIFICATIONS</i>	<i>RESULTS</i>
Turbidity	19 to 21 NTU	20.4 NTU

The expiration date is Oct 2017

Formazin and StablCal® solutions provided by Hach are not NIST traceable because the NIST does not carry turbidity standards. However, the use of Formazin and StablCal® as used in Hach method 8195 are accepted by the EPA as a primary standard to be used in the calibration of turbidity instruments.

Certified by _____

Scott Als
Analytical Services Chemist