

*Certificate of Analysis*

Page 1

COMMODITY: **StablCal® Standard, 1.0 NTU**COMMODITY NUMBER: **2659842**

MANUFACTURE DATE:

LOT NUMBER: **A9024****1/31/2019**

DATE OF ANALYSIS:

**2/4/2019**

---

<i><b>TEST</b></i>	<i><b>SPECIFICATIONS</b></i>	<i><b>RESULTS</b></i>
Turbidity	0.95 to 1.05 NTU	0.990 NTU

The expiration date is Jan 2021

Formazin and StablCal® solutions provided by Hach are not NIST traceable because the NIST does not carry turbidity standards. However, the use of Formazin and StablCal® as used in Hach method 8195 are accepted by the EPA as a primary standard to be used in the calibration of turbidity instruments.

Certified by \_\_\_\_\_

Scott Als  
Analytical Services Chemist