

*Certificate of Analysis*

Page 1

COMMODITY: **Standard 2, 5.0 mg/L as F**COMMODITY NUMBER: **2797111**

MANUFACTURE DATE:

LOT NUMBER: **A6321****11/14/2016**

DATE OF ANALYSIS:

11/14/2016

<i>TEST</i>	<i>SPECIFICATIONS</i>	<i>RESULTS</i>
Fluoride concentration by SPADNS method.	4.9 to 5.1 ppm	5.10 ppm

The expiration date is Nov 2021

The item 2797111 is traceable to NIST standards SRM 3183 Fluoride 1 mg/mL LOT 050721.

Certified by _____

Scott Als
Analytical Services Chemist