

***Certificate of Analysis***

Page 1

COMMODITY: **STABLCAL|sup|TS|sup0 FORMAZIN STANDARD 7500 NTU**
COMMODITY NUMBER: **2584201** MANUFACTURE DATE: **9/15/2021** DATE OF ANALYSIS: **9/15/2021**
LOT NUMBER: **A0313**

TEST	SPECIFICATIONS	RESULTS
Formazine Concentration	7125 to 7875 NTU	7521.5 NTU

The expiration date is Sep 2022

APROVADO
Responsável: <u>RENATO</u>
Padrão: <u>MRI - 7500 NTU</u>
Data: <u>13/10/2021</u>
Validade: <u>15/09/2022</u>

Certified by

Scott Als
Analytical Services Chemist