

*Certificate of Analysis*

Page 1

COMMODITY: **10 NTU Verification Standard**COMMODITY NUMBER: **2961801**

MANUFACTURE DATE:

DATE OF ANALYSIS:

LOT NUMBER: **A9088****4/4/2019****4/11/2019**

<i>TEST</i>	<i>SPECIFICATIONS</i>	<i>RESULTS</i>
Turbidity	9.5 to 10.5 NTU	10.00 NTU

The expiration date is Jul 2020

Formazin and StablCal® solutions provided by Hach are not NIST traceable because the NIST does not carry turbidity standards. However, the use of Formazin and StablCal® as used in Hach method 8195 are accepted by the EPA as a primary standard to be used in the calibration of turbidity instruments.

Certified by _____

Scott Als
Analytical Services Chemist