

*Certificate of Analysis*

Page 1

COMMODITY: **StablCal|sup|TS|sup0 Formazin Standard 4000 NTU**COMMODITY NUMBER: **246102**

MANUFACTURE DATE:

DATE OF ANALYSIS:

LOT NUMBER: **A2230****9/15/2022****9/22/2022**

TEST	SPECIFICATIONS	RESULTS
Turbidity	3800 to 4200 NTU	3908.0 NTU

The expiration date is Dec 2023

Formazin and StablCal® solutions provided by Hach are not NIST traceable because the NIS does not carry turbidity standards. However, the use of Formazin and StablCal® as use in Hach method 8195 are accepted by the EPA as a primary standard to be used in the calibration of turbidity instruments.

<b>APROVADO</b>
Responsável: <u>RENATO</u>
Padrão: <u>2461.02</u>
Data: <u>03/12/2022</u>
Validade: <u>31/12/2023</u>

Certified by

Scott Als  
Analytical Services Chemist