



## **Certificate of Analysis**

Product: DR/Check Secondary Standards Kit

Product Number: 27639-00 Lot Number: A3260 Expiration Date: Sep 2015

Product Number: 27639-00			Lot Number. Adads		
	Wavelength	Blank A3241	STD 1 (ABS) A3231	STD 2 (ABS) A3231	STD 3 (ABS) A3231
Г					
١					*
١	420 nm	0.000	<u>0.578</u> +/- 0.050	<u>1.170</u> +/- 0.100	<u>1.761</u> +/- 0.150
١	520 nm	0.000	0.600 +/- 0.050	<u>1.209</u> +/- 0.100	<u>1.816</u> +/- 0.150
	560 nm	0.000	<u>0.588</u> +/- 0.050	<u>1.184</u> +/- 0.100	<u>1.776</u> +/- 0.150
١	610 nm	0.000	<u>0.608</u> +/- 0.050	1.219 +/- 0.100	<u>1.826</u> +/- 0.150

NOTE: Transfer the control values to the enclosed certificate label and keep the label with your instrument for reference. For example, the test values for using a DR/890 Colorimeter at 560 nm would be:

0.588, 1.184, and 1.776 Absorbance for Standard 1, Standard 2, and Standard 3 respectively

File this certificate of Analysis for safe keeping. Certified at 25°C

Certified by:	Swe Cles	for and on the behalf of Hach Company