



Certificate of Analysis

1.19814.0500 Fluoride standard solution traceable to SRM from NIST NaF in H₂O 1000
mg/l F Certipur®
Batch HC85388314

Batch Values

Concentration	β (F ⁻)	1002	mg/l
---------------	---------------------	------	------

Determination method: Titration with Lanthannitrate solution.
(traceable to NIST SRM 682)
Accuracy of the method: ± 2 mg/l

Date of release (DD.MM.YYYY) 07.03.2018
Minimum shelf life (DD.MM.YYYY) 28.02.2021

Ayfer Yildirim

Responsible laboratory manager quality control

This document has been produced electronically and is valid without a signature.