

P.O.Box 389 Loveland, CO 80539 (970) 669-3050

Certificate of Analysis This is a Component of 2971205 lot A10032

Page 1

COMMODITY: STABLCAL | sup | TS | sup 0 FORMAZIN STANDARD

20 NTU

COMMODITY NUMBER: 2684801 MANUFACTURE DATE:

DATE OF ANALYSIS:

LOT NUMBER: A10012

4/26/2022

4/5/2022

TEST

SPECIFICATIONS

RESULTS

Turbidity

19 to 21 NTU

20.5 NTU

The expiration date is Apr 2023

Formazin and StablCal® solutions provided by Hach are not NIST traceable because the NIS does not carry turbidity standards. However, the use of Formazin and StablCal® as use in Hach method 8195 are accepted by the EPA as a primary standard to be used in the calibration of turbidity instruments.

APROVADO

Responsável: RENAMO

Padrão: Z6848.01

Data: OZ/06/Z0ZZ

Validade: 30/04/Z0Z3

Certified by .

Scott Als

Analytical Services Chemist

all als