

HACH COMPANY



An ISO 9001 Certified Company

P.O.Box 389
Loveland, CO 80539
(970) 669-3050

Certificate of Analysis

Page 1

COMMODITY: **STABLCAL® FORMAZIN STANDARD 200 NTU**
COMMODITY NUMBER: **2660401** MANUFACTURE DATE: **6/24/2020**
LOT NUMBER: **A0175**

DATE OF ANALYSIS: **6/29/2020**

<i>TEST</i>	<i>SPECIFICATIONS</i>	<i>RESULTS</i>
Turbidity	190 to 210 NTU	203.0 NTU

The expiration date is Sep 2021

Formazin and StablCal® solutions provided by Hach are not NIST traceable because the NIST does not carry turbidity standards. However, the use of Formazin and StablCal® as used in Hach method 8195 are accepted by the EPA as a primary standard to be used in the calibration of turbidity instruments.

APROVADO
Responsável: <u>Wellington</u>
Padrão: <u>2660401</u>
Data: <u>24/06/2020</u>
Validade: <u>09/2021</u>

Certified by _____

A handwritten signature in cursive script that reads "Scott Als".

Scott Als
Analytical Services Chemist