

***Certificate of Analysis***

Page 1

COMMODITY: **StablCal[®] Certified Standard 0.50**COMMODITY NUMBER: **2698053-BR** MANUFACTURE DATE:LOT NUMBER: **A0105****4/15/2020**

DATE OF ANALYSIS:

4/22/2020

| <i>TEST</i> | <i>SPECIFICATIONS</i> | <i>RESULTS</i> |
|----------------------------|------------------------------|-----------------------|
| Turbidity | 0.47 to 0.55 NTU | 0.527 NTU |
| Average Turbidity in mNTU | 470 to 550 mNTU | 527.0 mNTU |
| Standard Deviation in mNTU | 0 to 25 mNTU | 15.0 mNTU |
| Standard Deviation | 0 to 0.025 NTU | 0.0150 NTU |

The expiration date is Apr 2022

Certified by

Scott Als
Analytical Services Chemist