

Certificate of Analysis List

For request number 794194

Catalog Number Entered	Lot Number Entered	Related Catalog Number	Related Lot Code	Description
2660542	4259	N/A	N/A	StablCal [®] Standard, 800 NTU

Total Enclosures: 1

*Certificate of Analysis*

Page 1

COMMODITY: **StablCal® Standard, 800 NTU**COMMODITY NUMBER: **26605-42**

MANUFACTURE DATE:

DATE OF ANALYSIS:

LOT NUMBER: **A4259****9/18/2014****9/22/2014**

<i>TEST</i>	<i>SPECIFICATIONS</i>	<i>RESULTS</i>
Turbidity	760 to 840 NTU	808.5 NTU

The expiration date is Sep 2016

Formazin and StablCal® solutions provided by Hach are not NIST traceable because the NIST does not carry turbidity standards. However, the use of Formazin and StablCal® as used in Hach method 8195 are accepted by the EPA as a primary standard to be used in the calibration of turbidity instruments.

Certified by _____

Scott Als
Analytical Services Chemist