

*Certificate of Analysis*

Page 1

COMMODITY: **STABLCAL<sup>sup</sup> | TS<sup>sup</sup> | 0 FORMAZIN STANDARD 1000 NTU**  
COMMODITY NUMBER: **2660601** MANUFACTURE DATE: DATE OF ANALYSIS:  
LOT NUMBER: **A2264** **9/29/2022** **10/24/2022**

TEST	SPECIFICATIONS	RESULTS
Turbidity	950 to 1050 NTU	1015.0 NTU

The expiration date is Dec 2023

Formazin and StablCal® solutions provided by Hach are not NIST traceable because the NIS does not carry turbidity standards. However, the use of Formazin and StablCal® as use in Hach method 8195 are accepted by the EPA as a primary standard to be used in the calibration of turbidity instruments.

<b>APROVADO</b>	
Responsável:	RENATO
Padrão:	26606.01
Data:	03/12/2022
Validade:	31/12/2023

Certified by

Scott Als  
Analytical Services Chemist