

HACH COMPANY



An ISO 9001 Certified Company

P.O.Box 389  
Loveland, CO 80539  
(970) 669-3050

*Certificate of Analysis*

Page 1

COMMODITY: **Fluoride Standard Solution 1.00 mg/l as F**

COMMODITY NUMBER: **29153**

MANUFACTURE DATE:

DATE OF ANALYSIS:

LOT NUMBER: **A5240**

**8/31/2015**

**8/31/2015**

---

<i><b>TEST</b></i>	<i><b>SPECIFICATIONS</b></i>	<i><b>RESULTS</b></i>
Fluoride concentration by an electrode method.	0.98 to 1.02 ppm	1.000 ppm
Fluoride concentration by SPADNS method.	0.98 to 1.02 ppm	1.000 ppm

The expiration date is Aug 2020

The item 29153 is traceable to NIST standards SRM 3183 Fluoride 1 mg/mL LOT 050721.

Certified by \_\_\_\_\_

Scott Als  
Analytical Services Chemist