

HACH COMPANY



An ISO 9001 Certified Company

P.O.Box 389
Loveland, CO 80539
(970) 669-3050

Certificate of Analysis

Page 1

COMMODITY: **STABLCAL® FORMAZIN STANDARD 20 NTU**

COMMODITY NUMBER: **2684801** MANUFACTURE DATE:

LOT NUMBER: **A7345**

12/12/2017

DATE OF ANALYSIS:

12/29/2017

<i>TEST</i>	<i>SPECIFICATIONS</i>	<i>RESULTS</i>
Turbidity	19 to 21 NTU	20.4 NTU

The expiration date is Mar 2019

Formazin and StablCal® solutions provided by Hach are not NIST traceable because the NIST does not carry turbidity standards. However, the use of Formazin and StablCal® as used in Hach method 8195 are accepted by the EPA as a primary standard to be used in the calibration of turbidity instruments.

Certified by _____

Scott Als
Analytical Services Chemist