

HACH COMPANY



An ISO 9001 Certified Company

P.O.Box 389
Loveland, CO 80539
(970) 669-3050

Certificate of Analysis

This is a Component of 2971205-BR lot A2180

Page 1

COMMODITY: **10 NTU Verification Standard**

COMMODITY NUMBER: **2961801**

MANUFACTURE DATE:

DATE OF ANALYSIS:

LOT NUMBER: **A2171**

7/5/2022

7/11/2022

<i>TEST</i>	<i>SPECIFICATIONS</i>	<i>RESULTS</i>
Turbidity	9.5 to 10.5 NTU	10.20 NTU

The expiration date is Oct 2023

Formazin and StablCal® solutions provided by Hach are not NIST traceable because the NIST does not carry turbidity standards. However, the use of Formazin and StablCal® as used in Hach method 8195 are accepted by the EPA as a primary standard to be used in the calibration of turbidity instruments.

APROVADO
Responsável: <u>RENATO</u>
Padrão: <u>29618.01</u>
Data: <u>13/10/2022</u>
Validade: <u>30/10/2023</u>

Certified by _____

Scott Als

Scott Als
Analytical Services Chemist