

HACH COMPANY



An ISO 9001 Certified Company

P.O.Box 389
Loveland, CO 80539
(970) 669-3050

Certificate of Analysis

Page 1

COMMODITY: **StablCal® Formazin Standard 4000 NTU**

COMMODITY NUMBER: **246102**

MANUFACTURE DATE:

DATE OF ANALYSIS:

LOT NUMBER: **A0167**

6/22/2020

6/26/2020

<i>TEST</i>	<i>SPECIFICATIONS</i>	<i>RESULTS</i>
Turbidity	3800 to 4200 NTU	3996.0 NTU

The expiration date is Sep 2021

Formazin and StablCal® solutions provided by Hach are not NIST traceable because the NIST does not carry turbidity standards. However, the use of Formazin and StablCal® as used in Hach method 8195 are accepted by the EPA as a primary standard to be used in the calibration of turbidity instruments.

APROVADO
Responsável: <u>Wellington</u>
Padrão: <u>246102</u>
Data: <u>22/06/2020</u>
Validade: <u>09/2021</u>

Certified by _____

Scott Als
Analytical Services Chemist