

***Certificate of Analysis***

Page 1

COMMODITY: **StablCal|sup|TS|sup0 Standard, 1.0 NTU**COMMODITY NUMBER: **2659849**

MANUFACTURE DATE:

DATE OF ANALYSIS:

LOT NUMBER: **A2034****2/26/2022****3/1/2022**

TEST	SPECIFICATIONS	RESULTS
Turbidity	0.95 to 1.05 NTU	1.032 NTU

The expiration date is Feb 2024

Formazin and StablCal® solutions provided by Hach are not NIST traceable because the NIS does not carry turbidity standards. However, the use of Formazin and StablCal® as use in Hach method 8195 are accepted by the EPA as a primary standard to be used in the calibration of turbidity instruments.

APROVADO
Responsável: <u>RENATO</u>
Padrão: <u>26598.49</u>
Data: <u>03/12/2022</u>
Validade: <u>28/02/2024</u>

Certified by

Scott Als
Analytical Services Chemist