

HACH COMPANY



An ISO 9001 Certified Company

P.O.Box 389
Loveland, CO 80539
(970) 669-3050

Certificate of Analysis

This is a Component of 2971205-BR lot A2180

Page 1

COMMODITY: **STABLCAL|sup|TS|sup0 FORMAZIN STANDARD 800 NTU**

COMMODITY NUMBER: **2660501**

MANUFACTURE DATE:

DATE OF ANALYSIS:

LOT NUMBER: **A2127**

5/19/2022

5/26/2022

<i>TEST</i>	<i>SPECIFICATIONS</i>	<i>RESULTS</i>
Turbidity	760 to 840 NTU	821.0 NTU

The expiration date is Aug 2023

Formazin and StablCal® solutions provided by Hach are not NIST traceable because the NIST does not carry turbidity standards. However, the use of Formazin and StablCal® as used in Hach method 8195 are accepted by the EPA as a primary standard to be used in the calibration of turbidity instruments.

APROVADO
Responsável: <u>RENATO</u>
Padrão: <u>26605.01</u>
Data: <u>13/10/2022</u>
Validade: <u>30/08/2023</u>

Certified by _____

A handwritten signature in cursive script that reads "Scott Als".

Scott Als
Analytical Services Chemist