

HACH COMPANY



An ISO 9001 Certified Company

P.O.Box 389
Loveland, CO 80539
(970) 669-3050

Certificate of Analysis

This is a Component of 2971205-BR lot A1027

Page 1

COMMODITY: **STABLCAL® FORMAZIN STANDARD 20 NTU**

COMMODITY NUMBER: **2684801**

MANUFACTURE DATE:

DATE OF ANALYSIS:

LOT NUMBER: **A1028**

2/3/2021

2/8/2021

<i>TEST</i>	<i>SPECIFICATIONS</i>	<i>RESULTS</i>
Turbidity	19 to 21 NTU	20.5 NTU

The expiration date is May 2022

Formazin and StablCal® solutions provided by Hach are not NIST traceable because the NIST does not carry turbidity standards. However, the use of Formazin and StablCal® as used in Hach method 8195 are accepted by the EPA as a primary standard to be used in the calibration of turbidity instruments.

APROVADO
Responsável: <u>Wellington</u>
Padrão: <u>2684801-CJS</u>
Data: <u>06/07/2021</u>
Validade: <u>30/05/2022</u>

Certified by _____

Scott Als

Scott Als
Analytical Services Chemist