



An ISO 9001 Certified Company

Certificate of Analysis**This is a Component of 2659505-BR lot A1355**

Page 1

COMMODITY: **StablCal|sup|TS|sup0 Formazin Standard 4000 NTU**COMMODITY NUMBER: **246102**

MANUFACTURE DATE:

DATE OF ANALYSIS:

LOT NUMBER: **A1270****10/19/2021****10/25/2021**

<i>TEST</i>	<i>SPECIFICATIONS</i>	<i>RESULTS</i>
Turbidity	3800 to 4200 NTU	4007.0 NTU

The expiration date is Jan 2023

Formazin and StablCal® solutions provided by Hach are not NIST traceable because the NIS does not carry turbidity standards. However, the use of Formazin and StablCal® as use in Hach method 8195 are accepted by the EPA as a primary standard to be used in the calibration of turbidity instruments.

APROVADO
Responsável: <u>RENATO</u>
Padrão: <u>2461.02</u>
Data: <u>02/06/2022</u>
Validade: <u>30/01/2023</u>

Certified by

Scott Als
Analytical Services Chemist