



An ISO 9001 Certified Company

P.O.Box 389
Loveland, CO 80539
(970) 669-3050**Certificate of Analysis****This is a Component of 2971205 lot A0183**

Page 1

COMMODITY: **10 NTU Verification Standard**COMMODITY NUMBER: **2961801**

MANUFACTURE DATE:

DATE OF ANALYSIS:

LOT NUMBER: **A0164****6/17/2020****6/22/2020**

<i>TEST</i>	<i>SPECIFICATIONS</i>	<i>RESULTS</i>
Turbidity	9.5 to 10.5 NTU	10.00 NTU

The expiration date is Sep 2021

Formazin and StablCal® solutions provided by Hach are not NIST traceable because the NIS does not carry turbidity standards. However, the use of Formazin and StablCal® as use in Hach method 8195 are accepted by the EPA as a primary standard to be used in the calibration of turbidity instruments.

APROVADO
Responsável: <u>RENATO</u>
Padrão: <u>29618.01 - CT4</u>
Data: <u>21/09/2020</u>
Validade: <u>30/09/2021</u>

Certified by

Scott Als
Analytical Services Chemist