

Certificate of Analysis List

For request number 1109710

Catalog Number Entered	Lot Number Entered	Related Catalog Number	Related Lot Code	Description
246102-BR NTU	8323	2659505-BR	9018	StablCal® Formazin Standard 4000
2584201-BR NTU	9007	2659505-BR	9018	STABLCAL® FORMAZIN STANDARD 7500
2659701-BR	8289	2659505-BR	9018	<0.1 NTU Calibration Solution
2660101-BR NTU	8253	2659505-BR	9018	STABLCAL® FORMAZIN STANDARD 20
2660401-BR NTU	8257	2659505-BR	9018	STABLCAL® FORMAZIN STANDARD 200
2660601-BR NTU	8240	2659505-BR	9018	STABLCAL® FORMAZIN STANDARD 1000

Total Enclosures: 6

HACH COMPANY



An ISO 9001 Certified Company

P.O.Box 389
Loveland, CO 80539
(970) 669-3050

Certificate of Analysis

This is a Component of 2659505-BR lot A9018

Page 1

COMMODITY: **StablCal® Formazin Standard 4000 NTU**

COMMODITY NUMBER: **246102-BR** MANUFACTURE DATE:

DATE OF ANALYSIS:

LOT NUMBER: **A8323**

12/5/2018

12/10/2018

TEST

SPECIFICATIONS

RESULTS

Turbidity

3800 to 4200 NTU

3997.4 NTU

The expiration date is Mar 2020

Certified by _____

A handwritten signature in black ink that reads "Scott Als".

Scott Als
Analytical Services Chemist

HACH COMPANY



An ISO 9001 Certified Company

P.O.Box 389
Loveland, CO 80539
(970) 669-3050

Certificate of Analysis

This is a Component of 2659505-BR lot A9018

Page 1

COMMODITY: STABLCAL® FORMAZIN STANDARD 7500 NTU

COMMODITY NUMBER: 2584201-BR MANUFACTURE DATE:

LOT NUMBER: A9007

DATE OF ANALYSIS:

1/9/2019

1/14/2019

TEST

SPECIFICATIONS

RESULTS

Formazine Concentration

7125 to 7875 NTU

7463.3 NTU

The expiration date is Apr 2020

Certified by _____

A handwritten signature in black ink that reads "Scott Als".

Scott Als
Analytical Services Chemist

HACH COMPANY



An ISO 9001 Certified Company

P.O.Box 389
Loveland, CO 80539
(970) 669-3050

Certificate of Analysis

This is a Component of 2659505-BR lot A9018

Page 1

COMMODITY: <0.1 NTU Calibration Solution

COMMODITY NUMBER: 2659701-BR MANUFACTURE DATE:

LOT NUMBER: A8289

DATE OF ANALYSIS:

10/18/2018

10/26/2018

TEST

SPECIFICATIONS

RESULTS

Turbidity

0 to 0.1 NTU

0.04 NTU

The expiration date is Jan 2020

Certified by _____

A handwritten signature in black ink that reads "Scott Als".

Scott Als
Analytical Services Chemist

HACH COMPANY



An ISO 9001 Certified Company

P.O.Box 389
Loveland, CO 80539
(970) 669-3050

Certificate of Analysis

This is a Component of 2659505-BR lot A9018

Page 1

COMMODITY: STABLCAL® FORMAZIN STANDARD 20 NTU

COMMODITY NUMBER: 2660101-BR MANUFACTURE DATE:

DATE OF ANALYSIS:

LOT NUMBER: A8253

9/11/2018

9/18/2018

TEST

SPECIFICATIONS

RESULTS

Turbidity

19 to 21 NTU

20.5 NTU

The expiration date is Dec 2019

Certified by _____

A handwritten signature in black ink that reads "Scott Als".

Scott Als
Analytical Services Chemist

HACH COMPANY



An ISO 9001 Certified Company

P.O.Box 389
Loveland, CO 80539
(970) 669-3050

Certificate of Analysis

This is a Component of 2659505-BR lot A9018

Page 1

COMMODITY: STABLCAL® FORMAZIN STANDARD 200 NTU

COMMODITY NUMBER: 2660401-BR MANUFACTURE DATE:

DATE OF ANALYSIS:

LOT NUMBER: A8257

9/18/2018

9/24/2018

TEST

SPECIFICATIONS

RESULTS

Turbidity

190 to 210 NTU

204.3 NTU

The expiration date is Dec 2019

Certified by _____

A handwritten signature in black ink that reads "Scott Als".

Scott Als
Analytical Services Chemist

HACH COMPANY



An ISO 9001 Certified Company

P.O.Box 389
Loveland, CO 80539
(970) 669-3050

Certificate of Analysis

This is a Component of 2659505-BR lot A9018

Page 1

COMMODITY: STABLCAL® FORMAZIN STANDARD 1000 NTU

COMMODITY NUMBER: 2660601-BR MANUFACTURE DATE:

DATE OF ANALYSIS:

LOT NUMBER: A8240

8/31/2018

9/4/2018

TEST

SPECIFICATIONS

RESULTS

Turbidity

950 to 1050 NTU

1009.7 NTU

The expiration date is Nov 2019

Certified by _____

A handwritten signature in black ink that reads "Scott Als".

Scott Als
Analytical Services Chemist