



An ISO 9001 Certified Company

Certificate of Analysis**This is a Component of 2971205 lot A0183**

Page 1

COMMODITY: **STABLCAL® FORMAZIN STANDARD 20 NTU**COMMODITY NUMBER: **2684801**

MANUFACTURE DATE:

DATE OF ANALYSIS:

LOT NUMBER: **A0167****6/22/2020****6/26/2020**

TEST	SPECIFICATIONS	RESULTS
Turbidity	19 to 21 NTU	20.4 NTU

The expiration date is Sep 2021

Formazin and StablCal® solutions provided by Hach are not NIST traceable because the NIS does not carry turbidity standards. However, the use of Formazin and StablCal® as use in Hach method 8195 are accepted by the EPA as a primary standard to be used in the calibration of turbidity instruments.

APROVADO
Responsável: <u>RENATO</u>
Padrão: <u>26848.01 - CJ4</u>
Data: <u>21/09/2020</u>
Validade: <u>30/09/2021</u>

Certified by

Scott Als
Analytical Services Chemist