

Certificate of Analysis List

For request number 828013

Catalog Number Entered	Lot Number Entered	Related Catalog Number	Related Lot Code	Description
2684701	6175	N/A	N/A	<0.1 NTU Calibration Solution

Total Enclosures: 1



An ISO 9001 Certified Company

P.O.Box 389
Loveland, CO 80539
(970) 669-3050***Certificate of Analysis***

Page 1

COMMODITY: <0.1 NTU Calibration Solution

COMMODITY NUMBER: 26847-01

MANUFACTURE DATE:

DATE OF ANALYSIS:

LOT NUMBER: A6175

6/27/2016

7/5/2016

<i>TEST</i>	<i>SPECIFICATIONS</i>	<i>RESULTS</i>
Turbidity	0 to 0.1 NTU	0.06 NTU

The expiration date is Sep 2017

Formazin and StablCal® solutions provided by Hach are not NIST traceable because the NIST does not carry turbidity standards. However, the use of Formazin and StablCal® as used in Hach method 8195 are accepted by the EPA as a primary standard to be used in the calibration of turbidity instruments.

Certified by _____

Scott Als
Analytical Services Chemist