

*Certificate of Analysis*

Page 1

COMMODITY: **STABLCAL® FORMAZIN STANDARD 20 NTU**
COMMODITY NUMBER: **2684801** MANUFACTURE DATE:
LOT NUMBER: **A9088** **4/4/2019**

DATE OF ANALYSIS:
4/11/2019

<i>TEST</i>	<i>SPECIFICATIONS</i>	<i>RESULTS</i>
Turbidity	19 to 21 NTU	20.3 NTU

The expiration date is Jul 2020

Formazin and StablCal® solutions provided by Hach are not NIST traceable because the NIST does not carry turbidity standards. However, the use of Formazin and StablCal® as used in Hach method 8195 are accepted by the EPA as a primary standard to be used in the calibration of turbidity instruments.

Certified by _____

Scott Als
Analytical Services Chemist