



An ISO 9001 Certified Company

Certificate of Analysis***This is a Component of 2659505-BR lot A1355***

Page 1

COMMODITY: **STABLCAL | sup | TS | sup0 FORMAZIN STANDARD 7500 NTU**
COMMODITY NUMBER: **2584201** MANUFACTURE DATE: DATE OF ANALYSIS:
LOT NUMBER: **A1305** 11/3/2021 11/10/2021

TEST	SPECIFICATIONS	RESULTS
Formazine Concentration	7125 to 7875 NTU	7752.7 NTU

The expiration date is Feb 2023

APROVADO
Responsável: <u>RENATO</u>
Padrão: <u>25842.01</u>
Data: <u>02/06/2022</u>
Validade: <u>28/02/2023</u>

Certified by

Scott Als
Analytical Services Chemist