

HACH COMPANY



An ISO 9001 Certified Company

P.O.Box 389
Loveland, CO 80539
(970) 669-3050

Certificate of Analysis

Page 1

COMMODITY: **StablCal® Standard, 1.0 NTU**

COMMODITY NUMBER: **2659842**

MANUFACTURE DATE:

DATE OF ANALYSIS:

LOT NUMBER: **A7096**

4/13/2017

4/17/2017

| <i>TEST</i> | <i>SPECIFICATIONS</i> | <i>RESULTS</i> |
|--------------------|------------------------------|-----------------------|
| Turbidity | 0.95 to 1.05 NTU | 1.020 NTU |

The expiration date is Apr 2019

Formazin and StablCal® solutions provided by Hach are not NIST traceable because the NIST does not carry turbidity standards. However, the use of Formazin and StablCal® as used in Hach method 8195 are accepted by the EPA as a primary standard to be used in the calibration of turbidity instruments.

Certified by _____

Scott Als
Analytical Services Chemist