

Certificate of Analysis List

For request number 804571

Catalog Number Entered	Lot Number Entered	Related Catalog Number	Related Lot Code	Description
29153 as F	3070	N/A	N/A	Fluoride Standard Solution 1.00 mg/l

Total Enclosures: 1

*Certificate of Analysis*

Page 1

COMMODITY: **Fluoride Standard Solution 1.00 mg/l as F**COMMODITY NUMBER: **291-53**

MANUFACTURE DATE:

DATE OF ANALYSIS:

LOT NUMBER: **A3070****3/8/2013****3/11/2013**

<i>TEST</i>	<i>SPECIFICATIONS</i>	<i>RESULTS</i>
Fluoride concentration by an electrode method.	0.98 to 1.02 ppm	1.000 ppm
Fluoride concentration by SPADNS method.	0.98 to 1.02 ppm	0.980 ppm

The expiration date is Mar 2018

The item 29153 is traceable to NIST standards SRM 3183 Fluoride 1 mg/mL LOT 050721.

Certified by

Scott Als
Analytical Services Chemist